

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121823\
 Data File : PD080146.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Dec 2023 10:50
 Operator : AR\AJ
 Sample : INDA341
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA3220

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 17 21:00:25 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD121723CLP-TCLP.M
 Quant Title : GC Extractables
 QLast Update : Sat Dec 16 02:51:51 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.567	2.800	29277527	48895748	20.536	19.815
27)	SA Decachlor...	9.112	7.941	73490929	108.1E6	41.244	39.884
Target Compounds							
2)	A alpha-BHC	4.023	3.302	47789378	76774957	19.695	19.562
3)	MA gamma-BHC...	4.357	3.632	46102646	72183249	19.955	19.637
4)	MA Heptachlor	4.949	3.974	48471704	71203025	20.234	19.677
9)	A Endosulfan I	6.106	5.126	41722576	61680697	20.411	19.621
13)	MA Dieldrin	6.381	5.390	86461915	137.9E6	40.208	39.744
14)	MA Endrin	6.611	5.665	72809173	119.6E6	40.639	39.890
16)	A 4,4'-DDD	6.742	5.812	64716366	105.0E6	40.379	39.755
17)	MA 4,4'-DDT	7.057	6.063	69324699	109.7E6	40.287	39.470
20)	A Methoxychlor	7.531	6.638	183.9E6	286.3E6	205.485	206.617

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121823\
Data File : PD080146.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Dec 2023 10:50
Operator : AR\AJ
Sample : INDA341
Misc :
ALS Vial : 10 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 17 21:00:25 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD121723CLP-TCLP.M
Quant Title : GC Extractables
QLast Update : Sat Dec 16 02:51:51 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

