

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121924\
 Data File : PD087196.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Dec 2024 15:34
 Operator : AR\AJ
 Sample : INDB401
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDB4048

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 20 05:17:43 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD121924CLP-TCLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Dec 20 05:16:38 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.555	2.883	83632735	447.3E6	39.814	38.361
27)	SA Decachlor...	9.081	8.084	229.3E6	957.6E6	77.142	76.009
Target Compounds							
5)	MB Aldrin	5.277	4.375	156.0E6	623.9E6	41.841	38.744
6)	B beta-BHC	4.519	4.029	66995160	288.3E6	39.419	38.147
7)	B delta-BHC	4.768	4.266	162.9E6	650.9E6	42.976	39.243
8)	B Heptachlo...	5.697	4.879	140.8E6	574.0E6	40.940	38.254
10)	B trans-Chl...	5.952	5.133	139.5E6	602.9E6	41.414	38.685
11)	B cis-Chlor...	6.033	5.197	139.7E6	581.6E6	40.938	38.535
12)	B 4,4'-DDE	6.201	5.383	272.5E6	1119.6E6	85.324	78.054
15)	B Endosulfa...	6.791	6.088	240.9E6	993.7E6	82.850	76.356
18)	B Endrin al...	6.920	6.266	183.1E6	768.1E6	80.304	75.927
19)	B Endosulfa...	7.154	6.490	229.3E6	970.5E6	81.826	76.492
21)	B Endrin ke...	7.636	6.999	254.1E6	1068.7E6	81.834	76.414


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD121924\
Data File : PD087196.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Dec 2024 15:34
Operator : AR\AJ
Sample : INDB401
Misc :
ALS Vial : 17 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 20 05:17:43 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD121924CLP-TCLP.M
Quant Title : GC Extractables
QLast Update : Fri Dec 20 05:16:38 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.555 min
Delta R.T.: 0.000 min
Response: 83632735
Conc: 39.81 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDB4048




#1 Tetrachloro-m-xylene

R.T.: 2.883 min
Delta R.T.: 0.000 min
Response: 447313204
Conc: 38.36 ng/ml




#5 Aldrin

R.T.: 5.277 min
Delta R.T.: 0.000 min
Response: 155965186
Conc: 41.84 ng/ml



#5 Aldrin


R.T.: 4.375 min
Delta R.T.: 0.000 min
Response: 623875014
Conc: 38.74 ng/ml



#6 beta-BHC


R.T.: 4.519 min
Delta R.T.: 0.000 min
Response: 66995160
Conc: 39.42 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDB4048




#6 beta-BHC

R.T.: 4.029 min
Delta R.T.: 0.000 min
Response: 288256912
Conc: 38.15 ng/ml




#7 delta-BHC

R.T.: 4.768 min
Delta R.T.: 0.000 min
Response: 162929492
Conc: 42.98 ng/ml



#7 delta-BHC


R.T.: 4.266 min
Delta R.T.: 0.000 min
Response: 650878896
Conc: 39.24 ng/ml



#8 Heptachlor epoxide


R.T.: 5.697 min
Delta R.T.: -0.001 min
Response: 140773647
Conc: 40.94 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDB4048




#8 Heptachlor epoxide

R.T.: 4.879 min
Delta R.T.: 0.000 min
Response: 573964128
Conc: 38.25 ng/ml




#10 trans-Chlordane

R.T.: 5.952 min
Delta R.T.: 0.000 min
Response: 139456993
Conc: 41.41 ng/ml



#10 trans-Chlordane


R.T.: 5.133 min
Delta R.T.: 0.000 min
Response: 602857335
Conc: 38.69 ng/ml



#11 cis-Chlordane


R.T.: 6.033 min
Delta R.T.: -0.001 min
Response: 139684566
Conc: 40.94 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDB4048




#11 cis-Chlordane

R.T.: 5.197 min
Delta R.T.: 0.000 min
Response: 581554327
Conc: 38.54 ng/ml




#12 4,4'-DDE

R.T.: 6.201 min
Delta R.T.: 0.000 min
Response: 272475195
Conc: 85.32 ng/ml



#12 4,4'-DDE


R.T.: 5.383 min
Delta R.T.: 0.000 min
Response: 1119551179
Conc: 78.05 ng/ml



#15 Endosulfan II


R.T.: 6.791 min
Delta R.T.: 0.000 min
Response: 240895650
Conc: 82.85 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDB4048




#15 Endosulfan II

R.T.: 6.088 min
Delta R.T.: 0.000 min
Response: 993747971
Conc: 76.36 ng/ml




#18 Endrin aldehyde

R.T.: 6.920 min
Delta R.T.: 0.000 min
Response: 183110144
Conc: 80.30 ng/ml



#18 Endrin aldehyde


R.T.: 6.266 min
Delta R.T.: 0.000 min
Response: 768125197
Conc: 75.93 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.154 min
Delta R.T.: 0.000 min
Response: 229266678
Conc: 81.83 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDB4048




#19 Endosulfan Sulfate

R.T.: 6.490 min
Delta R.T.: 0.000 min
Response: 970464756
Conc: 76.49 ng/ml




#21 Endrin ketone

R.T.: 7.636 min
Delta R.T.: 0.000 min
Response: 254146856
Conc: 81.83 ng/ml



#21 Endrin ketone

R.T.: 6.999 min
Delta R.T.: 0.000 min
Response: 1068730915
Conc: 76.41 ng/ml



#27 Decachlorobiphenyl


R.T.: 9.081 min

Delta R.T.: -0.002 min

Response: 229261545

Conc: 77.14 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDB4048



#27 Decachlorobiphenyl

R.T.: 8.084 min

Delta R.T.: 0.000 min

Response: 957634174

Conc: 76.01 ng/ml