

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122022\
 Data File : PD073370.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Dec 2022 17:13
 Operator : AR\AJ
 Sample : N6129-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 C3

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 12/21/2022
 Supervised By : Ankita Jodhani 12/21/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 20 20:23:53 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD121622.M
 Quant Title : GC Extractables
 QLast Update : Fri Dec 16 15:46:21 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	2.694	3.455	20900073	799.2E6	9.972	8.127
28)	SA Decachlor...	7.821	8.901	27511546	198.5E6	11.363	11.258m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122022\
Data File : PD073370.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Dec 2022 17:13
Operator : AR\AJ
Sample : N6129-01
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
C3


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 12/21/2022
Supervised By : Ankita Jodhani 12/21/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 20 20:23:53 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD121622.M
Quant Title : GC Extractables
QLast Update : Fri Dec 16 15:46:21 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene

R.T.: 2.694 min
Delta R.T.: -0.002 min
Response: 20900073
Conc: 9.97 ng/ml

Instrument :
ECD_D
ClientSampleId :
C3


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 12/21/2022
Supervised By :Ankita Jodhani 12/21/2022




#1 Tetrachloro-m-xylene

R.T.: 3.455 min
Delta R.T.: -0.004 min
Response: 799241144
Conc: 8.13 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.821 min
Delta R.T.: -0.005 min
Response: 27511546
Conc: 11.36 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.901 min
Delta R.T.: -0.010 min
Response: 198457489
Conc: 11.26 ng/ml m