

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122120\
 Data File : PD060804.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Dec 2020 18:18
 Operator : DD\AJ
 Sample : INDA343
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA343

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 22 01:15:00 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120420CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Dec 04 03:58:39 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.501	4.022	20438129	88862621	17.494	22.592 #
27)	SA Decachlor...	8.290	9.109	47120820	128.8E6	31.608	39.795 #
Target Compounds							
2)	A alpha-BHC	3.942	4.413	29280398	131.1E6	17.387	22.208 #
3)	MA gamma-BHC...	4.228	4.698	27990792	118.5E6	17.354	22.098 #
4)	MA Heptachlor	4.534	5.210	29131195	117.3E6	17.348	22.296 #
9)	A Endosulfan I	5.572	6.262	25244856	94764397	17.375	22.139 #
13)	MA Dieldrin	5.818	6.520	54671756	202.8E6	34.426	43.409 #
14)	MA Endrin	6.076	6.738	42807723	146.5E6	33.085	39.733
16)	A 4,4'-DDD	6.200	6.857	43951962	158.2E6	34.207	42.924 #
17)	MA 4,4'-DDT	6.443	7.157	42868887	149.5E6	32.915	41.677 #
20)	A Methoxychlor	6.991	7.614	113.4E6	358.4E6	166.771	198.651

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122120\
Data File : PD060804.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Dec 2020 18:18
Operator : DD\AJ
Sample : INDA343
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 22 01:15:00 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120420CLP.M
Quant Title : GC Extractables
QLast Update : Fri Dec 04 03:58:39 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

