

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122123\
Data File : PD080221.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Dec 2023 15:05
Operator : AR\AJ
Sample : 05794-17DL2 50X
Misc :
ALS Vial : 16 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
BGRG0DL2

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 12/20/2023
Supervised By :Ankita Jodhani 12/21/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 20 01:07:08 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD121623CLP.M
Quant Title : GC Extractables
QLast Update : Mon Dec 18 00:44:33 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(14) Endrin (MA)

6.608min 18.770 ng/ml

response 29399286

(14) Endrin #2 (MA)

5.673min -13.854 ng/ml

response -36450396

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122123\
Data File : PD080221.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Dec 2023 15:05
Operator : AR\AJ
Sample : 05794-17DL2 50X
Misc :
ALS Vial : 16 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
BGRG0DL2

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 12/20/2023
Supervised By :Ankita Jodhani 12/21/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 20 01:07:08 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD121623CLP.M
Quant Title : GC Extractables
QLast Update : Mon Dec 18 00:44:33 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(14) Endrin (MA)

6.608min 18.770 ng/ml

response 29399286

(14) Endrin #2 (MA)

5.633min 22.884 ng/ml m

response 60209046

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122123\
Data File : PD080221.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Dec 2023 15:05
Operator : AR\AJ
Sample : 05794-17DL2 50X
Misc :
ALS Vial : 16 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
BGRG0DL2

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 12/20/2023
Supervised By :Ankita Jodhani 12/21/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 20 01:07:08 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD121623CLP.M
Quant Title : GC Extractables
QLast Update : Mon Dec 18 00:44:33 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(17) 4,4'-DDT (MA)

7.059min 27.609 ng/ml

response 40798950

(17) 4,4'-DDT #2 (MA)

6.063min 49.398 ng/ml

response 119107932

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122123\
Data File : PD080221.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Dec 2023 15:05
Operator : AR\AJ
Sample : 05794-17DL2 50X
Misc :
ALS Vial : 16 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
BGRG0DL2

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 12/20/2023
Supervised By :Ankita Jodhani 12/21/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 20 01:07:08 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD121623CLP.M
Quant Title : GC Extractables
QLast Update : Mon Dec 18 00:44:33 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(17) 4,4'-DDT (MA)

7.058min 31.590 ng/ml m

response 46681880

(17) 4,4'-DDT #2 (MA)

6.063min 49.398 ng/ml

response 119107932