

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122223\
Data File : PD080290.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Dec 2023 21:09
Operator : AR\AJ
Sample : PIBLK127
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :

ECD_D

LabSampleId :

PIBLK127


Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 12/21/2023

Supervised By :Ankita Jodhani 12/22/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 21 00:05:54 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD121623CLP.M
Quant Title : GC Extractables
QLast Update : Mon Dec 18 00:44:33 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(27) Decachlorobiphenyl (SA)

9.111min 24.139 ng/ml

response 38809914

(27) Decachlorobiphenyl #2 (SA)

7.940min 23.585 ng/ml

response 56608807

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122223\
Data File : PD080290.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Dec 2023 21:09
Operator : AR\AJ
Sample : PIBLK127
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :

ECD_D

LabSampleId :

PIBLK127


Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 12/21/2023

Supervised By :Ankita Jodhani 12/22/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 21 00:05:54 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD121623CLP.M
Quant Title : GC Extractables
QLast Update : Mon Dec 18 00:44:33 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(27) Decachlorobiphenyl (SA)

9.110min 21.818 ng/ml m

response 35077935

(27) Decachlorobiphenyl #2 (SA)

7.940min 23.585 ng/ml

response 56608807

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122223\
 Data File : PD080290.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Dec 2023 21:09
 Operator : AR\AJ
 Sample : PIBLK127
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_D
 LabSampleID :
 PIBLK127

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 12/21/2023
 Supervised By :Ankita Jodhani 12/22/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 21 00:05:54 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD121623CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Dec 18 00:44:33 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.566	2.800	27680197	49264702	21.680	22.156
27) SA Decachlor...	9.110	7.940	35077935	56608807	21.818m	23.585

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122223\
Data File : PD080290.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Dec 2023 21:09
Operator : AR\AJ
Sample : PIBLK127
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
LabSampleId :
PIBLK127

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 12/21/2023
Supervised By :Ankita Jodhani 12/22/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 21 00:05:54 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD121623CLP.M
Quant Title : GC Extractables
QLast Update : Mon Dec 18 00:44:33 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

