

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122619\
 Data File : PD056786.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Dec 2019 07:59
 Operator : SG\AJ
 Sample : PIBLK42
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 27 10:12:57 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120219CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Dec 03 08:51:13 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds						
1) SA Tetrachlo...	3.286	3.960	18341358	31912383	22.146	26.544
27) SA Decachlor...	7.958	8.993	35526828	48538767	39.344	40.343
Target Compounds						
6) B beta-BHC	0.000	4.845f	0	1694800	N.D.	2.057
8) B Heptachlo...	0.000	5.816	0	12460706	N.D.	8.043
14) MA Endrin	0.000	6.649	0	2774246	N.D.	2.192
20) A Methoxychlor	0.000	7.464f	0	1422224	N.D.	2.348
23) Toxaphene-2	0.000	6.649	0	2774246	N.D.	201.027


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122619\
Data File : PD056786.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Dec 2019 07:59
Operator : SG\AJ
Sample : PIBLK42
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 27 10:12:57 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD120219CLP.M
Quant Title : GC Extractables
QLast Update : Tue Dec 03 08:51:13 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.286 min
Delta R.T.: -0.002 min
Response: 18341358
Conc: 22.15 ng/ml

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.960 min
Delta R.T.: -0.003 min
Response: 31912383
Conc: 26.54 ng/ml




#6 beta-BHC

R.T.: 0.000 min
Exp R.T. : 4.240 min
Response: 0
Conc: N.D.



#6 beta-BHC


R.T.: 4.845 min
Delta R.T.: 0.044 min
Response: 1694800
Conc: 2.06 ng/ml



#8 Heptachlor epoxide


R.T.: 0.000 min
Exp R.T.: 4.953 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#8 Heptachlor epoxide

R.T.: 5.816 min
Delta R.T.: -0.010 min
Response: 12460706
Conc: 8.04 ng/ml




#14 Endrin

R.T.: 0.000 min
Exp R.T.: 5.778 min
Response: 0
Conc: N.D.



#14 Endrin


R.T.: 6.649 min
Delta R.T.: -0.006 min
Response: 2774246
Conc: 2.19 ng/ml



#20 Methoxychlor


R.T.: 0.000 min
Exp R.T.: 6.696 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#20 Methoxychlor

R.T.: 7.464 min
Delta R.T.: -0.069 min
Response: 1422224
Conc: 2.35 ng/ml




#23 Toxaphene-2

R.T.: 0.000 min
Exp R.T.: 5.644 min
Response: 0
Conc: N.D.



#23 Toxaphene-2


R.T.: 6.649 min
Delta R.T.: 0.021 min
Response: 2774246
Conc: 201.03 ng/ml



#27 Decachlorobiphenyl

R.T.: 7.958 min
Delta R.T.: -0.007 min
Response: 35526828
Conc: 39.34 ng/ml

Instrument :
ECD_D
ClientSampleId :



#27 Decachlorobiphenyl

R.T.: 8.993 min
Delta R.T.: -0.010 min
Response: 48538767
Conc: 40.34 ng/ml