

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122624\
Data File : PD087350.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Dec 2024 16:04
Operator : AR\AJ
Sample : P5334-02
Misc :
ALS Vial : 14 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
E2905

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 12/27/2024
Supervised By :Ankita Jodhani 12/27/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 26 21:49:52 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD122024CLP.M
Quant Title : GC Extractables
QLast Update : Fri Dec 20 23:32:53 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(13) Dieldrin (MA)

6.354min 1.120 ng/ml

response 4049490

(13) Dieldrin #2 (MA)

5.516min 3.254 ng/ml

response 49808940

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122624\
Data File : PD087350.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Dec 2024 16:04
Operator : AR\AJ
Sample : P5334-02
Misc :
ALS Vial : 14 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
E2905

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 12/27/2024
Supervised By :Ankita Jodhani 12/27/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 26 21:49:52 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD122024CLP.M
Quant Title : GC Extractables
QLast Update : Fri Dec 20 23:32:53 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(13) Dieldrin (MA)

6.353min 1.583 ng/ml m
response 5725577

(13) Dieldrin #2 (MA)

5.514min 4.357 ng/ml m
response 66708428

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122624\
Data File : PD087350.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Dec 2024 16:04
Operator : AR\AJ
Sample : P5334-02
Misc :
ALS Vial : 14 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
E2905

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 12/27/2024
Supervised By :Ankita Jodhani 12/27/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 26 21:49:52 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD122024CLP.M
Quant Title : GC Extractables
QLast Update : Fri Dec 20 23:32:53 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(14) Endrin (MA)

6.580min 2.387 ng/ml
response 7130874

(14) Endrin #2 (MA)

5.797min -1.123 ng/ml
response -15695413

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122624\
Data File : PD087350.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Dec 2024 16:04
Operator : AR\AJ
Sample : P5334-02
Misc :
ALS Vial : 14 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
E2905

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 12/27/2024
Supervised By :Ankita Jodhani 12/27/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 26 21:49:52 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD122024CLP.M
Quant Title : GC Extractables
QLast Update : Fri Dec 20 23:32:53 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(14) Endrin (MA)
6.580min 2.387 ng/ml
response 7130874

(14) Endrin #2 (MA)
5.796min 1.295 ng/ml m
response 18087679

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122624\
Data File : PD087350.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Dec 2024 16:04
Operator : AR\AJ
Sample : P5334-02
Misc :
ALS Vial : 14 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
E2905

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 12/27/2024
Supervised By :Ankita Jodhani 12/27/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 26 21:49:52 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD122024CLP.M
Quant Title : GC Extractables
QLast Update : Fri Dec 20 23:32:53 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(15) Endosulfan II (B)

6.801min -0.619 ng/ml

response -1848885

(15) Endosulfan II #2 (B)

6.076min 2.960 ng/ml

response 40185472

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122624\
Data File : PD087350.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Dec 2024 16:04
Operator : AR\AJ
Sample : P5334-02
Misc :
ALS Vial : 14 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
E2905

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 12/27/2024
Supervised By :Ankita Jodhani 12/27/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 26 21:49:52 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD122024CLP.M
Quant Title : GC Extractables
QLast Update : Fri Dec 20 23:32:53 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(15) Endosulfan II (B)
6.799min 1.686 ng/ml m
response 5038761

(15) Endosulfan II #2 (B)
6.076min 2.960 ng/ml
response 40185472

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122624\
Data File : PD087350.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Dec 2024 16:04
Operator : AR\AJ
Sample : P5334-02
Misc :
ALS Vial : 14 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
E2905

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 12/27/2024
Supervised By :Ankita Jodhani 12/27/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 26 21:49:52 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD122024CLP.M
Quant Title : GC Extractables
QLast Update : Fri Dec 20 23:32:53 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(17) 4,4'-DDT (MA)

0.000min 0.000 ng/ml

response 0

(17) 4,4'-DDT #2 (MA)

6.186min 8.274 ng/ml

response 104311872

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122624\
Data File : PD087350.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Dec 2024 16:04
Operator : AR\AJ
Sample : P5334-02
Misc :
ALS Vial : 14 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
E2905

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 12/27/2024
Supervised By :Ankita Jodhani 12/27/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 26 21:49:52 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD122024CLP.M
Quant Title : GC Extractables
QLast Update : Fri Dec 20 23:32:53 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(17) 4,4'-DDT (MA)

7.001min 7.416 ng/ml m

response 19882972

(17) 4,4'-DDT #2 (MA)

6.185min 9.359 ng/ml m

response 117993464

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122624\
Data File : PD087350.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Dec 2024 16:04
Operator : AR\AJ
Sample : P5334-02
Misc :
ALS Vial : 14 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
E2905

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 12/27/2024
Supervised By :Ankita Jodhani 12/27/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 26 21:49:52 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD122024CLP.M
Quant Title : GC Extractables
QLast Update : Fri Dec 20 23:32:53 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(19) Endosulfan Sulfate (B)

7.150min 1.387 ng/ml

response 4019364

(19) Endosulfan Sulfate #2 (B)

6.490min 10.567 ng/ml

response 139705984

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122624\
Data File : PD087350.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Dec 2024 16:04
Operator : AR\AJ
Sample : P5334-02
Misc :
ALS Vial : 14 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
E2905

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 12/27/2024
Supervised By :Ankita Jodhani 12/27/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 26 21:49:52 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD122024CLP.M
Quant Title : GC Extractables
QLast Update : Fri Dec 20 23:32:53 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(19) Endosulfan Sulfate (B)

7.149min 2.507 ng/ml m

response 7266466

(19) Endosulfan Sulfate #2 (B)

6.488min 7.701 ng/ml m

response 101810790