

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122718\
 Data File : PD050882.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Dec 2018 13:45
 Operator : AJ\SJ
 Sample : INDA401
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 27 14:06:28 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD122718CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Dec 27 14:01:23 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.227	4.045	62128212	1019.0E6	40.152	35.511
27)	SA Decachlor...	7.828	9.150	82763205	1355.2E6	68.450	67.113
Target Compounds							
2)	A alpha-BHC	3.647	4.437	100.4E6	1575.3E6	45.257	37.116
3)	MA gamma-BHC...	3.915	4.724	94699111	1442.8E6	42.976	36.636
4)	MA Heptachlor	4.193	5.238	100.1E6	1437.5E6	42.856	36.090
9)	A Endosulfan I	5.173	6.290	87671352	1111.9E6	41.392	35.175
13)	MA Dieldrin	5.409	6.548	203.1E6	2273.6E6	88.895	72.599
14)	MA Endrin	5.656	6.767	165.3E6	1722.2E6	84.651	71.888
16)	A 4,4'-DDD	5.797	6.886	145.3E6	1769.1E6	86.987	73.711
17)	MA 4,4'-DDT	6.028	7.186	154.0E6	1664.5E6	88.361	78.086
20)	A Methoxychlor	6.574	7.644	365.3E6	3698.5E6	404.285	358.620


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122718\
Data File : PD050882.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 27 Dec 2018 13:45
Operator : AJ\SJ
Sample : INDA401
Misc :
ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 27 14:06:28 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD122718CLP.M
Quant Title : GC Extractables
QLast Update : Thu Dec 27 14:01:23 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.227 min
Delta R.T.: 0.000 min
Response: 62128212
Conc: 40.15 ng/ml

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.045 min
Delta R.T.: 0.000 min
Response: 1019015844
Conc: 35.51 ng/ml




#2 alpha-BHC

R.T.: 3.647 min
Delta R.T.: 0.000 min
Response: 100390194
Conc: 45.26 ng/ml



#2 alpha-BHC


R.T.: 4.437 min
Delta R.T.: 0.000 min
Response: 1575281880
Conc: 37.12 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 3.915 min
Delta R.T.: 0.000 min
Response: 94699111
Conc: 42.98 ng/ml

Instrument :
ECD_D
ClientSampleId :




#3 gamma-BHC (Lindane)

R.T.: 4.724 min
Delta R.T.: 0.000 min
Response: 1442847953
Conc: 36.64 ng/ml




#4 Heptachlor

R.T.: 4.193 min
Delta R.T.: 0.000 min
Response: 100077422
Conc: 42.86 ng/ml



#4 Heptachlor


R.T.: 5.238 min
Delta R.T.: 0.000 min
Response: 1437523306
Conc: 36.09 ng/ml



#9 Endosulfan I


R.T.: 5.173 min
Delta R.T.: 0.000 min
Response: 87671352
Conc: 41.39 ng/ml

Instrument :
ECD_D
ClientSampleId :




#9 Endosulfan I

R.T.: 6.290 min
Delta R.T.: 0.000 min
Response: 1111858783
Conc: 35.18 ng/ml




#13 Dieldrin

R.T.: 5.409 min
Delta R.T.: 0.000 min
Response: 203101099
Conc: 88.89 ng/ml



#13 Dieldrin


R.T.: 6.548 min
Delta R.T.: 0.000 min
Response: 2273550535
Conc: 72.60 ng/ml



#14 Endrin


R.T.: 5.656 min
Delta R.T.: 0.000 min
Response: 165291471
Conc: 84.65 ng/ml

Instrument :
ECD_D
ClientSampleId :




#14 Endrin

R.T.: 6.767 min
Delta R.T.: 0.000 min
Response: 1722166012
Conc: 71.89 ng/ml




#16 4,4'-DDD

R.T.: 5.797 min
Delta R.T.: 0.000 min
Response: 145308428
Conc: 86.99 ng/ml



#16 4,4'-DDD


R.T.: 6.886 min
Delta R.T.: 0.000 min
Response: 1769148495
Conc: 73.71 ng/ml



#17 4,4'-DDT


R.T.: 6.028 min
Delta R.T.: 0.000 min
Response: 153956427
Conc: 88.36 ng/ml

Instrument :
ECD_D
ClientSampleId :




#17 4,4'-DDT

R.T.: 7.186 min
Delta R.T.: 0.000 min
Response: 1664459978
Conc: 78.09 ng/ml




#20 Methoxychlor

R.T.: 6.574 min
Delta R.T.: 0.000 min
Response: 365314725
Conc: 404.28 ng/ml



#20 Methoxychlor


R.T.: 7.644 min
Delta R.T.: 0.000 min
Response: 3698499662
Conc: 358.62 ng/ml



#27 Decachlorobiphenyl

R.T.: 7.828 min
Delta R.T.: 0.000 min
Response: 82763205
Conc: 68.45 ng/ml

Instrument :
ECD_D
ClientSampleId :



#27 Decachlorobiphenyl

R.T.: 9.150 min
Delta R.T.: -0.001 min
Response: 1355171863
Conc: 67.11 ng/ml