

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122722\
 Data File : PD073484.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Dec 2022 11:35
 Operator : AR\AJ
 Sample : RESC046
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 RESC049

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 28 05:06:46 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD122722CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Dec 28 03:07:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	2.693	3.455	51636782	1278.3E6	20.037	18.621
27)	SA Decachlor...	7.818	8.901	121.3E6	709.3E6	40.574	39.770
Target Compounds							
2)	A alpha-BHC	3.195	3.918	40122340	849.8E6	9.972	9.907
3)	MA gamma-BHC...	3.526	4.251	31253156	678.1E6	9.267	10.156
4)	MA Heptachlor	3.857	4.817	35678221	467.0E6	9.678	9.785
5)	MB Aldrin	4.136	5.155	33508355	392.2E6	9.038	8.859
6)	B beta-BHC	3.827	4.467	15029739	237.4E6	8.900	9.287
7)	B delta-BHC	4.056	4.710	34334202	417.0E6	8.717	8.157m
8)	B Heptachlo...	4.640	5.587	30615795	314.5E6	8.978	8.216
9)	A Endosulfan I	5.009	5.972	32503599	337.5E6	10.127	9.768
10)	B trans-Chl...	4.889	5.841	30889023	313.8E6	8.990	8.480
11)	B cis-Chlor...	4.951	5.916	31484183	325.6E6	8.999	8.657
12)	B 4,4'-DDE	5.140	6.090	61814358	622.1E6	17.881	17.290
13)	MA Dieldrin	5.275	6.248	71043311	710.4E6	20.027	19.408
14)	MA Endrin	5.551	6.480	62319839	620.5E6	19.965	20.339
15)	B Endosulfa...	5.849	6.710	55249813	476.8E6	18.350	17.275
16)	A 4,4'-DDD	5.697	6.615	56176902	527.4E6	20.213	19.807
17)	MA 4,4'-DDT	5.946	6.921	54965314	519.4E6	19.940	19.988
18)	B Endrin al...	6.031	6.841	47652240	403.8E6	18.895	17.803
19)	B Endosulfa...	6.253	7.074	55031055	456.5E6	18.807	17.994
20)	A Methoxychlor	6.529	7.402	158.2E6	1155.7E6	106.011	103.476
21)	B Endrin ke...	6.764	7.563	67242002	494.8E6	19.369	18.902

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

```
Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122722\
Data File : PD073484.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Dec 2022 11:35
Operator : AR\AJ
Sample : RESC046
Misc :
ALS Vial : 2 Sample Multiplier: 1
```

Integration File signal 1: autoint1.e

Integration File signal 2: autoint2.e

Quant Time: Dec 28 05:06:46 2022

Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD D\Method\PD122722CLP.M

Quant Title : GC Extractables

OLast Update : Wed Dec 28 03:07:05 2022

Response via : Initial Calibration

Integrator: ChemStation

Volume Inj. : 1 μ l

Signal #1 Phase : ZB-MR2

Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50um

