

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122723\
 Data File : PD080506.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Dec 2023 18:07
 Operator : AR\AJ
 Sample : 05986-03MS
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 C0AG1MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 26 20:34:10 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD121623CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Dec 21 13:52:23 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.573	2.796	28200408	46583370	22.087	20.950
2)	SA Decachlor...	9.118	7.939	42414042	56558580	26.381	23.564
Target Compounds							
2)	A alpha-BHC	4.030	3.299	122.9E6	196.8E6	59.620	57.728
3)	MA gamma-BHC...	4.363	3.628	114.6E6	179.9E6	57.698	55.892
4)	MA Heptachlor	4.956	3.970	111.6E6	169.6E6	53.441	52.948
5)	MB Aldrin	5.298	4.250	105.8E6	163.0E6	47.310	46.130
6)	B beta-BHC	4.557	3.924	47562243	74891882	45.821	43.617
7)	B delta-BHC	4.807	4.155	116.5E6	184.0E6	50.702	47.334
8)	B Heptachlo...	5.725	4.752	104.2E6	162.6E6	48.518	47.964
9)	A Endosulfan I	6.112	5.123	99084014	152.1E6	55.622	55.470
10)	B trans-Chl...	5.981	5.003	100.2E6	157.4E6	48.842	47.825
11)	B cis-Chlor...	6.061	5.067	102.9E6	159.8E6	49.012	47.528
12)	B 4,4'-DDE	6.232	5.256	190.4E6	299.9E6	98.555	95.475
13)	MA Dieldrin	6.387	5.387	217.0E6	340.4E6	116.609	113.597
14)	MA Endrin	6.617	5.663	172.5E6	289.3E6	110.127	109.974
15)	B Endosulfa...	6.836	5.955	184.1E6	290.3E6	98.181	100.126
16)	A 4,4'-DDD	6.748	5.809	159.8E6	260.5E6	115.183	113.279
17)	MA 4,4'-DDT	7.064	6.061	163.9E6	253.2E6	110.934	104.991
18)	B Endrin al...	6.965	6.135	136.1E6	211.8E6	91.221	91.868
19)	B Endosulfa...	7.200	6.357	176.4E6	272.7E6	96.189	97.223
20)	A Methoxychlor	7.538	6.636	418.4E6	616.4E6	528.412	489.652
21)	B Endrin ke...	7.686	6.863	195.9E6	319.9E6	100.682	99.978

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122723\
Data File : PD080506.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Dec 2023 18:07
Operator : AR\AJ
Sample : 05986-03MS
Misc :
ALS Vial : 14 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 26 20:34:10 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD121623CLP.M
Quant Title : GC Extractables
QLast Update : Thu Dec 21 13:52:23 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

