

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122823\
 Data File : PD080547.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Dec 2023 09:09
 Operator : AR\AJ
 Sample : PEM076
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEM088

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 12/28/2023
 Supervised By : Ankita Jodhani 12/28/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 27 20:25:15 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD121623CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Dec 21 13:52:23 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.567	2.798	29189321	49957524	22.861	22.467
27)	SA Decachlor...	9.115	7.940	39336053	53293711	24.466	22.203
Target Compounds							
2)	A alpha-BHC	4.024	3.301	23344394	39343654	11.324	11.540
3)	MA gamma-BHC...	4.357	3.630	23089075	37048582	11.627	11.509
6)	B beta-BHC	4.552	3.925	10933259	18128915	10.533	10.558
12)	B 4,4'-DDE	6.225	5.257	459401	968019	0.238m	0.308m#
14)	MA Endrin	6.614	5.664	77944727	134.9E6	49.764	51.263
16)	A 4,4'-DDD	6.744	5.811	4847033	8107352	3.494	3.526
17)	MA 4,4'-DDT	7.060	6.062	155.1E6	255.7E6	104.956	106.062
18)	B Endrin al...	6.962	6.137	3406438	6878535	2.284	2.983 #
20)	A Methoxychlor	7.534	6.636	200.7E6	324.3E6	253.527	257.651
21)	B Endrin ke...	7.683	6.864	6723212	13600672	3.455	4.250

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122823\
Data File : PD080547.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Dec 2023 09:09
Operator : AR\AJ
Sample : PEM076
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 27 20:25:15 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD121623CLP.M
Quant Title : GC Extractables
QLast Update : Thu Dec 21 13:52:23 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Instrument :
ECD_D
ClientSampled :
PEM088

Manual Integrations
APPROVED

Reviewed By: Abdul Miza 12/28/2023
Supervised By: Ankita Jodhani 12/28/2023