

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122923\
 Data File : PD080577.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Dec 2023 12:37
 Operator : AR\AJ
 Sample : PB158063BS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PLCS063

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 28 21:16:45 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD121623CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Dec 21 13:52:23 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.567	2.798	29614624	50496126	23.195	22.710
2)	SA Decachlor...	9.118	7.942	44190930	56688438	27.486	23.618
Target Compounds							
2)	A alpha-BHC	4.024	3.301	11124222	18070752	5.396	5.300
3)	MA gamma-BHC...	4.358	3.630	10585914	17007792	5.331	5.284
4)	MA Heptachlor	4.951	3.972	11908590	16611295	5.703	5.187
5)	MB Aldrin	5.295	4.252	11573235	17543619	5.174	4.966
6)	B beta-BHC	4.552	3.926	5355469	7737821	5.159	4.507
7)	B delta-BHC	4.801	4.156	11964445	16952055	5.209	4.360
8)	B Heptachlo...	5.721	4.754	11165842	16161045	5.198	4.769
9)	A Endosulfan I	6.109	5.124	10722180	15240380	6.019	5.559
10)	B trans-Chl...	5.978	5.005	10700064	15819954	5.214	4.806
11)	B cis-Chlor...	6.058	5.068	11066060	16192043	5.272	4.816
12)	B 4,4'-DDE	6.229	5.258	18997560	28521099	9.834	9.080
13)	MA Dieldrin	6.384	5.389	21231918	31456622	11.408	10.496
14)	MA Endrin	6.615	5.665	16210289	25622538	10.349	9.738
15)	B Endosulfa...	6.833	5.957	19810712	28669995	10.567	9.889
16)	A 4,4'-DDD	6.746	5.812	16190988	24998353	11.673	10.872
17)	MA 4,4'-DDT	7.062	6.063	15229883	22041726	10.306	9.142
18)	B Endrin al...	6.963	6.137	15682487	23004957	10.513	9.976
19)	B Endosulfa...	7.197	6.359	19308723	27339218	10.527	9.747
20)	A Methoxychlor	7.536	6.638	42098976	62653433	53.170	49.772
21)	B Endrin ke...	7.685	6.866	20697554	32387009	10.637	10.121

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD122923\
Data File : PD080577.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Dec 2023 12:37
Operator : AR\AJ
Sample : PB158063BS
Misc :
ALS Vial : 7 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 28 21:16:45 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD121623CLP.M
Quant Title : GC Extractables
QLast Update : Thu Dec 21 13:52:23 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

