

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102119\
Data File : PD055472.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2019 14:18
Operator : SG\AJ
Sample : INDA327
Misc :
ALS Vial : 10 Sample Multiplier: 1


Instrument :
ECD_D
LabSampleId :
INDA327

Manual Integrations
APPROVED

Ankita
10/23/2019 10:09:22 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 03:58:38 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101819CLP.M
Quant Title : GC Extractables
QLast Update : Sat Oct 19 00:42:03 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(9) Endosulfan I (A)
5.287min 21.246 ng/ml
response 20421754

(9) Endosulfan I #2 (A)
6.180min 20.125 ng/ml
response 32985772

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102119\
Data File : PD055472.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2019 14:18
Operator : SG\AJ
Sample : INDA327
Misc :
ALS Vial : 10 Sample Multiplier: 1


Instrument :
ECD_D
LabSampleId :
INDA327

Manual Integrations
APPROVED

Ankita
10/23/2019 10:09:22 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 03:58:38 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101819CLP.M
Quant Title : GC Extractables
QLast Update : Sat Oct 19 00:42:03 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(9) Endosulfan I (A)

5.287min 21.246 ng/ml

response 20421754

(9) Endosulfan I #2 (A)

6.178min 19.813 ng/ml m

response 32474771

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102119\
 Data File : PD055472.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2019 14:18
 Operator : SG\AJ
 Sample : INDA327
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_D
 LabSampleId :
 INDA327

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 22 03:58:38 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101819CLP.M
 Quant Title : GC Extractables
 Qlast Update : Sat Oct 19 00:42:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Manual Integrations
 APPROVED

Ankita
 10/23/2019 10:09:22 AM

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.291	3.962	19234083	27970748	21.573	20.775
27) SA Decachlor...	7.967	8.997	35311223	45268942	38.563	34.694
Target Compounds						
2) A alpha-BHC	3.718	4.349	27423437	40931904	21.837	21.040
3) MA gamma-BHC...	3.994	4.632	25212457	37916702	21.276	20.563
4) MA Heptachlor	4.279	5.138	20399033	37090776	22.418	19.890
9) A Endosulfan I	5.287	6.178	20421754	32474771	21.246	19.813m
13) MA Dieldrin	5.527	6.436	47636737	72046592	42.984	39.808
14) MA Endrin	5.780	6.652	33874663	47896518	37.819	34.291
16) A 4,4'-DDD	5.914	6.774	36904333	58669575	45.534	41.010
17) MA 4,4'-DDT	6.150	7.072	26652758	46844593	39.311	34.561
20) A Methoxychlor	6.698	7.529	74453426	133.9E6	204.167	173.649

AJ
 10/24/19

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.