

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111224\
 Data File : PD086576.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Nov 2024 05:06
 Operator : AR\AJ
 Sample : PB164822BS
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PLCS822

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 11/12/2024
 Supervised By :Ankita Jodhani 11/12/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 12 06:49:17 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102524CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 25 14:36:38 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.557	2.886	28039474	211.2E6	21.297	22.773
27)	SA Decachlor...	9.094	8.093	40632398	240.3E6	22.665	25.415
Target Compounds							
2)	A alpha-BHC	4.007	3.400	12983431	94311230	5.215	6.564 #
3)	MA gamma-BHC...	4.339	3.738	13201364	90828147	5.306	6.734 #
4)	MA Heptachlor	4.940	4.093	13625078	93205382	5.511	6.822m
5)	MB Aldrin	5.282f	4.379	13278951	90677797	5.479	6.684m
6)	B beta-BHC	4.523	4.032	6225399	41272722	5.606	6.583m
7)	B delta-BHC	4.772	4.271	13595462	88760704	5.252	6.326
8)	B Heptachlo...	5.702	4.884	12557869	83237117	5.495	6.638m
9)	A Endosulfan I	6.086	5.261	11996054	80408832	5.922	7.003
10)	B trans-Chl...	5.957	5.138	11914416	87931718	5.400	6.684m
11)	B cis-Chlor...	6.039	5.204	12607258	88351550	5.702	6.956
12)	B 4,4'-DDE	6.207	5.390	22401946	157.2E6	10.488	12.820
13)	MA Dieldrin	6.359	5.527	25230714	161.7E6	11.061	13.107
14)	MA Endrin	6.587	5.802	20572083	150.7E6	10.945	13.409m
15)	B Endosulfa...	6.798	6.095	20947595	145.4E6	10.449	13.178 #
16)	A 4,4'-DDD	6.717	5.945	18116171	135.2E6	11.180	13.770
17)	MA 4,4'-DDT	7.034	6.199	18140957	130.9E6	10.881	12.753
18)	B Endrin al...	6.927	6.273	17481289	112.0E6	11.719	13.493
19)	B Endosulfa...	7.161	6.497	21443907	143.7E6	11.633	13.633
20)	A Methoxychlor	7.506	6.772	54922976	330.6E6	58.360	69.544
21)	B Endrin ke...	7.643	7.007	24249402	158.6E6	11.767	13.862

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD111224\
Data File : PD086576.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2024 05:06
Operator : AR\AJ
Sample : PB164822BS
Misc :
ALS Vial : 45 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 06:49:17 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102524CLP.M
Quant Title : GC Extractables
QLast Update : Fri Oct 25 14:36:38 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Manual Integrations
APPROVED
Reviewed By: Abdul Mirza 11/12/2024
Supervised By: Ankita Jodhani 11/12/2024

Instrument :
ECD_D
Client Sampled :
PLCS822