

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042721\
 Data File : PD062486.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Apr 2021 13:23
 Operator : AR\AJ
 Sample : M2169-09DL 2X
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 CF-9DL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 27 15:45:43 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD042321.M
 Quant Title : GC Extractables
 QLast Update : Fri Apr 23 23:19:57 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.484	3.981	8405702	114.2E6	10.320	10.847
28)	SA Decachlor...	8.256	9.040	12409556	94751777	10.230	11.167
Target Compounds							
12)	B 4,4'-DDE	5.654	6.316	86151591	962.3E6	70.972	68.151
13)	MA Dieldrin	5.790	6.466	1918272	23394607	1.503	1.663
16)	A 4,4'-DDD	6.172	6.806	2469078	37040414	2.467	3.324 #
17)	MA 4,4'-DDT	6.414	7.104	76781711	780.7E6	75.190	78.286



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042721\
Data File : PD062486.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Apr 2021 13:23
Operator : AR\AJ
Sample : M2169-09DL 2X
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
CF-9DL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 27 15:45:43 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD042321.M
Quant Title : GC Extractables
QLast Update : Fri Apr 23 23:19:57 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.484 min
Delta R.T.: 0.000 min
Response: 8405702
Conc: 10.32 ng/ml

Instrument :
ECD_D
ClientSampleId :
CF-9DL




#1 Tetrachloro-m-xylene

R.T.: 3.981 min
Delta R.T.: 0.000 min
Response: 114235875
Conc: 10.85 ng/ml




#12 4,4'-DDE

R.T.: 5.654 min
Delta R.T.: -0.001 min
Response: 86151591
Conc: 70.97 ng/ml



#12 4,4'-DDE


R.T.: 6.316 min
Delta R.T.: -0.001 min
Response: 962270811
Conc: 68.15 ng/ml



#13 Dieldrin


R.T.: 5.790 min
Delta R.T.: -0.002 min
Response: 1918272
Conc: 1.50 ng/ml

Instrument :
ECD_D
ClientSampleId :
CF-9DL




#13 Dieldrin

R.T.: 6.466 min
Delta R.T.: -0.001 min
Response: 23394607
Conc: 1.66 ng/ml




#16 4,4'-DDD

R.T.: 6.172 min
Delta R.T.: -0.001 min
Response: 2469078
Conc: 2.47 ng/ml



#16 4,4'-DDD


R.T.: 6.806 min
Delta R.T.: -0.001 min
Response: 37040414
Conc: 3.32 ng/ml



#17 4,4'-DDT


R.T.: 6.414 min
Delta R.T.: 0.000 min
Response: 76781711
Conc: 75.19 ng/ml

Instrument :
ECD_D
ClientSampleId :
CF-9DL




#17 4,4'-DDT

R.T.: 7.104 min
Delta R.T.: -0.001 min
Response: 780680967
Conc: 78.29 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.256 min
Delta R.T.: -0.002 min
Response: 12409556
Conc: 10.23 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.040 min
Delta R.T.: -0.002 min
Response: 94751777
Conc: 11.17 ng/ml