

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042722\
 Data File : PD069257.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Apr 2022 10:53
 Operator : AR\AJ
 Sample : PEM220
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEM020

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 04/28/2022
 Supervised By :mohammad ahmed 04/28/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 27 21:32:10 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041422CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Apr 21 02:24:30 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.449	4.171	24856069	235.7E6	19.618	19.526
27)	SA Decachlor...	8.206	9.369	27826148	140.0E6	20.925	19.912
Target Compounds							
2)	A alpha-BHC	3.889	4.568	17593112	173.1E6	9.842	10.528
3)	MA gamma-BHC...	4.172	4.859	17877542	163.8E6	9.884	10.513
6)	B beta-BHC	4.423	5.023	8654482	74652719	11.173	11.129
12)	B 4,4'-DDE	5.611	6.550	491508	4668520	0.366m	0.429
14)	MA Endrin	6.002	6.928	73513243	438.7E6	52.540	45.107
16)	A 4,4'-DDD	6.130	7.041	1963483	16899629	1.610	1.805
17)	MA 4,4'-DDT	6.370	7.343	132.0E6	859.3E6	107.606	95.462
18)	B Endrin al...	6.444	7.257	583162	2457049	0.587	0.344m#
20)	A Methoxychlor	6.920	7.797	172.3E6	992.8E6	253.939	240.911
21)	B Endrin ke...	7.149	7.959	3522263	10479860	2.438	1.135 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD042722\
Data File : PD069257.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Apr 2022 10:53
Operator : AR\AJ
Sample : PEM220
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 27 21:32:10 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD041422CLP.M
Quant Title : GC Extractables
QLast Update : Thu Apr 21 02:24:30 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Reviewed By: Abdul Miza 04/28/2022
Supervised By: mohammad ahmed 04/28/2022

Manual Integrations
APPROVED

Instrument :
ECD_D
ClientSampled :
PEM020