

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050823\
 Data File : PD075330.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 May 2023 13:55
 Operator : AR\AJ
 Sample : INDA412
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA3235

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 09 04:17:21 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu May 04 02:14:07 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.551	2.772	49042110	23797091	21.502	22.562
27) SA Decachlor...	9.096	7.920	86317432	36300815	34.880	38.439
Target Compounds						
2) A alpha-BHC	4.007	3.274	78346356	33170583	20.373	21.414
3) MA gamma-BHC...	4.341	3.604	73602409	31045831	20.473	21.121
4) MA Heptachlor	4.933	3.945	74539578	29331722	19.995	20.074
9) A Endosulfan I	6.090	5.099	60062608	25189400	19.912	20.988
13) MA Dieldrin	6.366	5.364	131.9E6	55084812	38.584	40.373
14) MA Endrin	6.597	5.640	103.4E6	45313507	35.743	37.255
16) A 4,4'-DDD	6.730	5.788	102.5E6	43625704	42.398	44.005
17) MA 4,4'-DDT	7.046	6.039	84458845	35613123	33.275	35.335
20) A Methoxychlor	7.520	6.615	246.7E6	98496290	167.951	171.961

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050823\
Data File : PD075330.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 May 2023 13:55
Operator : AR\AJ
Sample : INDA412
Misc :
ALS Vial : 5 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 09 04:17:21 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

