

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050823\
 Data File : PD075347.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 May 2023 20:11
 Operator : AR\AJ
 Sample : 02609-16MSD
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 JREJ4MSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 09 04:22:16 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu May 04 02:14:07 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.549	2.771	46538446	22382797	20.404	21.221
2)	SA Decachlor...	9.092	7.918	67767536	28836983	27.384	30.535
Target Compounds							
2)	A alpha-BHC	4.006	3.273	184.7E6	76905364	48.034	49.648
3)	MA gamma-BHC...	4.339	3.603	168.5E6	69801686	46.866	47.486
4)	MA Heptachlor	4.930	3.944	167.8E6	65417213	45.021	44.770
5)	MB Aldrin	5.274	4.225	173.3E6	68339990	47.523	46.079
6)	B beta-BHC	4.535	3.900	69731763	30186243	42.481	41.456
7)	B delta-BHC	4.783	4.131	132.7E6	56048061	39.262	35.471
8)	B Heptachlo...	5.701	4.727	162.6E6	61701915	47.972	44.238
9)	A Endosulfan I	6.088	5.098	143.7E6	57769706	47.636	48.135
10)	B trans-Chl...	5.958	4.978	152.3E6	61391480	46.466	44.990
11)	B cis-Chlor...	6.037	5.042	155.2E6	65700044	45.540	47.077
12)	B 4,4'-DDE	6.209	5.231	301.3E6	117.4E6	92.275	91.333
13)	MA Dieldrin	6.363	5.363	335.8E6	131.7E6	98.192	96.540
14)	MA Endrin	6.594	5.638	275.9E6	114.3E6	95.398	93.977
15)	B Endosulfa...	6.813	5.933	275.4E6	108.3E6	85.521	84.641
16)	A 4,4'-DDD	6.727	5.786	252.3E6	100.2E6	104.396	101.049
17)	MA 4,4'-DDT	7.042	6.038	227.2E6	91462587	89.494	90.748
18)	B Endrin al...	6.943	6.112	193.0E6	77632177	81.588	76.449
19)	B Endosulfa...	7.178	6.335	238.7E6	97237280	86.775	83.607
20)	A Methoxychlor	7.517	6.614	623.8E6	239.2E6	424.631	417.636
21)	B Endrin ke...	7.665	6.842	280.6E6	118.7E6	92.558	90.039

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050823\
Data File : PD075347.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 May 2023 20:11
Operator : AR\AJ
Sample : 02609-16MSD
Misc :
ALS Vial : 20 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 09 04:22:16 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

