

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061021\
 Data File : PD063718.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Jun 2021 10:10
 Operator : AR\AJ
 Sample : M2620-06
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 11 03:58:23 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060321CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Jun 03 06:08:10 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds


6) B	beta-BHC	4.377f	4.855f	2990302	62627900	3.215	7.001 #
7) B	delta-BHC	0.000	5.007	0	23609078	N.D.	1.070
8) B	Heptachlo...	0.000	5.852	0	6548157	N.D.	0.329
16) A	4,4'-DDD	0.000	6.754f	0	7585694	N.D.	0.499
20) A	Methoxychlor	0.000	7.483f	0	2657221	N.D.	0.368
26)	Toxaphene-5	7.058	0.000	1476613	0	43.101	N.D. #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061021\
Data File : PD063718.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jun 2021 10:10
Operator : AR\AJ
Sample : M2620-06
Misc :
ALS Vial : 4 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 11 03:58:23 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060321CLP.M
Quant Title : GC Extractables
QLast Update : Thu Jun 03 06:08:10 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 0.000 min
Exp R.T.: 3.439 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T.: 3.970 min
Response: 0
Conc: N.D.




#6 beta-BHC

R.T.: 4.377 min
Delta R.T.: -0.033 min
Response: 2990302
Conc: 3.21 ng/ml



#6 beta-BHC


R.T.: 4.855 min
Delta R.T.: 0.041 min
Response: 62627900
Conc: 7.00 ng/ml



#7 delta-BHC


R.T.: 0.000 min
Exp R.T. : 4.613 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#7 delta-BHC

R.T.: 5.007 min
Delta R.T.: -0.024 min
Response: 23609078
Conc: 1.07 ng/ml




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T. : 5.150 min
Response: 0
Conc: N.D.



#8 Heptachlor epoxide


R.T.: 5.852 min
Delta R.T.: 0.011 min
Response: 6548157
Conc: 0.33 ng/ml



#16 4,4'-DDD


R.T.: 0.000 min
Exp R.T.: 6.115 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#16 4,4'-DDD

R.T.: 6.754 min
Delta R.T.: -0.039 min
Response: 7585694
Conc: 0.50 ng/ml




#20 Methoxychlor

R.T.: 0.000 min
Exp R.T.: 6.902 min
Response: 0
Conc: N.D.



#20 Methoxychlor


R.T.: 7.483 min
Delta R.T.: -0.067 min
Response: 2657221
Conc: 0.37 ng/ml



#26 Toxaphene-5


R.T.: 7.058 min
Delta R.T.: 0.029 min
Response: 1476613
Conc: 43.10 ng/ml

Instrument :
ECD_D
ClientSampleId :




#26 Toxaphene-5

R.T.: 0.000 min
Exp R.T. : 7.625 min
Response: 0
Conc: N.D.



#27 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 8.194 min
Response: 0
Conc: N.D.



#27 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 9.024 min
Response: 0
Conc: N.D.