

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061024\
Data File : PD083573.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jun 2024 11:16
Operator : AR\AJ
Sample : PEM111
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_D
LabSampleId :
PEM111

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 06/11/2024
Supervised By :Ankita Jodhani 06/11/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 10 22:04:29 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060724CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jun 07 04:03:14 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(16) 4,4'-DDD (A)

0.000min 0.000 ng/ml

response 0

(16) 4,4'-DDD #2 (A)

0.000min 0.000 ng/ml

response 0

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061024\
Data File : PD083573.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jun 2024 11:16
Operator : AR\AJ
Sample : PEM111
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :

ECD_D

LabSampleId :

PEM111



Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 06/11/2024

Supervised By :Ankita Jodhani 06/11/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 10 22:04:29 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060724CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jun 07 04:03:14 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(16) 4,4'-DDD (A)

6.746min 0.185 ng/ml m

response 271383

(16) 4,4'-DDD #2 (A)

5.775min 0.279 ng/ml m

response 1571558

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061024\
Data File : PD083573.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jun 2024 11:16
Operator : AR\AJ
Sample : PEM111
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :

ECD_D

LabSampleId :

PEM111



Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 06/11/2024

Supervised By :Ankita Jodhani 06/11/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 10 22:04:29 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060724CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jun 07 04:03:14 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(18) Endrin aldehyde (B)

6.964min 1.054 ng/ml

response 1406236

(18) Endrin aldehyde #2 (B)

0.000min 0.000 ng/ml

response 0

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061024\
Data File : PD083573.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jun 2024 11:16
Operator : AR\AJ
Sample : PEM111
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_D
LabSampleId :
PEM111

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 06/11/2024
Supervised By :Ankita Jodhani 06/11/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 10 22:04:29 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060724CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jun 07 04:03:14 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(18) Endrin aldehyde (B)
6.964min 1.054 ng/ml
response 1406236

(18) Endrin aldehyde #2 (B)
6.101min 1.513 ng/ml m
response 7246307

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061024\
Data File : PD083573.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jun 2024 11:16
Operator : AR\AJ
Sample : PEM111
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :

ECD_D

LabSampleId :

PEM111



Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 06/11/2024

Supervised By :Ankita Jodhani 06/11/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 10 22:04:29 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060724CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jun 07 04:03:14 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(27) Decachlorobiphenyl (SA)

9.122min 20.957 ng/ml

response 40751294

(27) Decachlorobiphenyl #2 (SA)

7.902min 20.401 ng/ml

response 123878406

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061024\
Data File : PD083573.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jun 2024 11:16
Operator : AR\AJ
Sample : PEM111
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_D
LabSampleId :
PEM111

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 06/11/2024
Supervised By :Ankita Jodhani 06/11/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 10 22:04:29 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060724CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jun 07 04:03:14 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(27) Decachlorobiphenyl (SA)

9.122min 20.957 ng/ml

response 40751294

(27) Decachlorobiphenyl #2 (SA)

7.900min 20.055 ng/ml m

response 121774156