Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061024\
Data File : PD083603.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2024 01:17
Operator : AR\AJ
Sample : P2659-09
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
BHC30

Manual Integrations APPROVED

Reviewed By :Abdul Mirza 06/12/2024
Supervised By :mohammad ahmed 06/13/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 11 09:34:57 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060724CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jun 07 04:03:14 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.561	2.771	10286420	52611872	7.880m	9.652m
27)	SA Decachlor...	9.105	7.896	31993344	133.8E6	16.453m	22.028m#
	Target Compounds						
9)	A Endosulfan I	6.138	5.116	47492973	161.4E6	24.286m	24.417m
10)	B trans-Chl...	5.964	4.967	10517244	21334711	5.165m	2.937m#
11)	B cis-Chlor...	6.055	5.030	31495452	36085092	15.124m	4.901m#
12)	B 4,4'-DDE	6.226	5.218	23305139	136.3E6	11.236	19.185m#
13)	MA Dieldrin	6.382	5.350	309.0E6	1118.9E6	149.452	153.588m
14)	MA Endrin	6.625	5.629	16980767	33123539	10.052m	5.122m#
16)	A 4,4'-DDD	6.734	5.775	18376956	20241067	12.541m	3.590m#
17)	MA 4,4'-DDT	7.060	6.022	45178015	189.9E6	29.990m	33.542m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061024\
Data File : PD083603.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2024 01:17
Operator : AR\AJ
Sample : P2659-09
Misc :
ALS Vial : 30 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 11 09:34:57 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060724CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jun 07 04:03:14 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm





Manual Integrations APPROVED
Reviewed By :Abdul Mirza 06/12/2024
Supervised By :mohammad ahmed 06/13/2024

Instrument :
ECD_D
ClientSampleId :
BHC30