

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061424\
 Data File : PD083697.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jun 2024 10:27
 Operator : AR\AJ
 Sample : P2659-20
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 BHC41

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 06/17/2024
 Supervised By : Ankita Jodhani 06/17/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 15 02:26:11 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060724CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 07 04:03:14 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.567	2.767	4793129	21322032	3.672m	3.912m
27)	SA Decachlor...	9.122	7.897	35328405	112.6E6	18.168m	18.551
Target Compounds							
10)	B trans-Chl...	5.980	4.965	3041223	7294528	1.494m	1.004m#
11)	B cis-Chlor...	6.063	5.027	8148564	14255695	3.913	1.936m#
12)	B 4,4'-DDE	6.233	5.216	15031442	75227588	7.247	10.586m#
13)	MA Dieldrin	6.388	5.348	26412337	141.5E6	12.774	19.420m#
16)	A 4,4'-DDD	6.743	5.772	2581675	4657193	1.762m	0.826m#
17)	MA 4,4'-DDT	7.068	6.022	29822086	104.5E6	19.796	18.467

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061424\
Data File : PD083697.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jun 2024 10:27
Operator : AR\AJ
Sample : P2659-20
Misc :
ALS Vial : 4 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 15 02:26:11 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060724CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jun 07 04:03:14 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

