

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD070122\  
 Data File : PD070658.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 01 Jul 2022 06:42  
 Operator : AR\AJ  
 Sample : INDA101  
 Misc :  
 ALS Vial : 37 Sample Multiplier: 1

Instrument :  
 ECD\_D  
 ClientSampleId :  
 INDA1047

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Jul 01 10:09:18 2022  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD070122CLP.M  
 Quant Title : GC Extractables  
 QLast Update : Fri Jul 01 10:09:11 2022  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 µl  
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
-----						
System Monitoring Compounds						
1) SA Tetrachlo...	3.487	4.126	12438235	139.1E6	4.918	4.985
27) SA Decachlor...	8.267	9.315	25770784	75496403	10.143	9.979
Target Compounds						
2) A alpha-BHC	3.929	4.530	16315112	191.7E6	4.729	4.905
3) MA gamma-BHC...	4.215	4.824	16611355	177.8E6	4.861	4.966
4) MA Heptachlor	4.517	5.339	16174095	168.4E6	4.795	4.965
9) A Endosulfan I	5.552	6.410	14403874	107.7E6	4.880	4.986
13) MA Dieldrin	5.796	6.671	30599138	208.9E6	9.510	9.922
14) MA Endrin	6.055	6.894	25476664	144.9E6	9.525	9.683
16) A 4,4'-DDD	6.182	7.009	24862241	162.7E6	9.555	10.026
17) MA 4,4'-DDT	6.424	7.309	23448191	134.4E6	9.405	9.739
20) A Methoxychlor	6.975	7.769	67517524	306.6E6	48.840	49.344
-----						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD070122\  
 Data File : PD070658.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 01 Jul 2022 06:42  
 Operator : AR\AJ  
 Sample : INDA101  
 Misc :  
 ALS Vial : 37 Sample Multiplier: 1

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Jul 01 10:09:18 2022  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD070122CLP.M  
 Quant Title : GC Extractables  
 QLast Update : Fri Jul 01 10:09:11 2022  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 µl  
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

