

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD070323\
 Data File : PD076286.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jul 2023 10:53
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 03 21:36:02 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061523.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 16 10:43:19 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.555	2.766	103.4E6	52796480	48.531	48.664
28)	SA Decachlor...	9.101	7.915	114.0E6	49494205	44.986	50.618
Target Compounds							
2)	A alpha-BHC	4.012	3.269	184.2E6	80658851	49.891	50.420
3)	MA gamma-BHC...	4.346	3.599	166.7E6	74446577	49.547	50.526
4)	MA Heptachlor	4.937	3.940	169.6E6	72924929	47.300	48.013
5)	MB Aldrin	5.280	4.220	168.9E6	74142539	48.211	49.720
6)	B beta-BHC	4.542	3.897	68372234	31521201	47.556	49.452
7)	B delta-BHC	4.791	4.127	171.2E6	75209859	51.497	51.905
8)	B Heptachlo...	5.707	4.723	152.9E6	67003680	47.455	48.778
9)	A Endosulfan I	6.095	5.094	143.7E6	62633948	47.894	49.001
10)	B gamma-Chl...	5.964	4.974	161.8E6	68376353	50.227	49.833
11)	B alpha-Chl...	6.043	5.038	156.1E6	67595465	48.284	49.276
12)	B 4,4'-DDE	6.217	5.228	141.3E6	61228510	49.455	54.856
13)	MA Dieldrin	6.370	5.359	151.0E6	67570948	48.319	50.083
14)	MA Endrin	6.601	5.635	119.1E6	56072484	44.522	46.075
15)	B Endosulfa...	6.820	5.930	120.0E6	54376900	46.446	50.403
16)	A 4,4'-DDD	6.734	5.783	119.0E6	52022595	52.308	59.031
17)	MA 4,4'-DDT	7.049	6.035	110.2E6	45749746	48.557	56.322
18)	B Endrin al...	6.951	6.109	93420505	42545068	46.557	49.157
19)	B Endosulfa...	7.186	6.332	121.4E6	55359609	48.191	51.230
20)	A Methoxychlor	7.525	6.610	58231097	23461931	44.987	50.691
21)	B Endrin ke...	7.672	6.839	141.9E6	64924439	50.923	53.438
22)	Mirex	8.148	7.022	98822099	50471202	46.124	48.149

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD070323\
Data File : PD076286.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jul 2023 10:53
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 03 21:36:02 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061523.M
Quant Title : GC Extractables
QLast Update : Fri Jun 16 10:43:19 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

