

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD071223\
 Data File : PD076562.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jul 2023 11:38
 Operator : AR\AJ
 Sample : 03414-16DL 5X
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 A46E6DL

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 07/13/2023
 Supervised By : Ankita Jodhani 07/13/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 12 20:38:22 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD070523CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Jul 05 14:30:19 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.537	2.765	442.4E6	2105541	204.045	2.046 #
27)	SA Decachlor...	9.074	7.907	17741510	6384515	6.994m	6.997
Target Compounds							
3)	MA gamma-BHC...	4.340	3.575	375.2E6	97958686	110.455m	68.068m#
7)	B delta-BHC	4.798	4.132	244.9E6	37880957	71.741m	26.643 #
11)	B cis-Chlor...	6.023	5.042	12194890	11723304	3.824	9.037m#
14)	MA Endrin	6.586	5.628	20793958	2575016	8.058m	2.272m#
16)	A 4,4'-DDD	6.713	5.776	33341354	8540885	14.602	10.572m#
17)	MA 4,4'-DDT	7.010	6.021	15682604	6848546	6.926m	8.217m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD071223\
Data File : PD076562.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jul 2023 11:38
Operator : AR\AJ
Sample : 03414-16DL 5X
Misc :
ALS Vial : 5 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 12 20:38:22 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD070523CLP.M
Quant Title : GC Extractables
QLast Update : Wed Jul 05 14:30:19 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Reviewed By: Abdul Miza 07/13/2023
Supervised By: Ankita Jodhani 07/13/2023

Manual Integrations
APPROVED

Instrument :
ECD_D
ClientSampled :
A46E6DL