

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081224\
Data File : PD084320.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2024 20:49
Operator : AR\AJ
Sample : P3171-08RX
Misc :
ALS Vial : 43 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
DCVX0RX

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :Ankita Jodhani 08/14/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 13 10:25:47 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081224CLP.M
Quant Title : GC Extractables
QLast Update : Mon Aug 12 16:40:36 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

3.541min 212.025 ng/ml

response 177756024

(1) Tetrachloro-m-xylene #2 (SA)

2.927min -16.678 ng/ml

response -87794733

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081224\
Data File : PD084320.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2024 20:49
Operator : AR\AJ
Sample : P3171-08RX
Misc :
ALS Vial : 43 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :


DCVX0RX

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :Ankita Jodhani 08/14/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 13 10:25:47 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081224CLP.M
Quant Title : GC Extractables
QLast Update : Mon Aug 12 16:40:36 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

3.586min 17.863 ng/ml m

response 14976140

(1) Tetrachloro-m-xylene #2 (SA)

2.926min 10.797 ng/ml m

response 56836266

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081224\
Data File : PD084320.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2024 20:49
Operator : AR\AJ
Sample : P3171-08RX
Misc :
ALS Vial : 43 Sample Multiplier: 1

Instrument :

ECD_D

Client SampleId :

DCVX0RX


Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :Ankita Jodhani 08/14/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 13 10:25:47 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081224CLP.M
Quant Title : GC Extractables
QLast Update : Mon Aug 12 16:40:36 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(7) delta-BHC (B)

4.797min 169.015 ng/ml

response 291875795

(7) delta-BHC #2 (B)

4.310min 54.724 ng/ml

response 459031821

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081224\
Data File : PD084320.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2024 20:49
Operator : AR\AJ
Sample : P3171-08RX
Misc :
ALS Vial : 43 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
DCVX0RX

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :Ankita Jodhani 08/14/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 13 10:25:47 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081224CLP.M
Quant Title : GC Extractables
QLast Update : Mon Aug 12 16:40:36 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(7) delta-BHC (B)
4.794min 129.039 ng/ml m
response 222840516

(7) delta-BHC #2 (B)
4.309min 47.294 ng/ml m
response 396703879

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081224\
Data File : PD084320.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2024 20:49
Operator : AR\AJ
Sample : P3171-08RX
Misc :
ALS Vial : 43 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
DCVX0RX

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :Ankita Jodhani 08/14/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 13 10:25:47 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081224CLP.M
Quant Title : GC Extractables
QLast Update : Mon Aug 12 16:40:36 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(8) Heptachlor epoxide (B)

0.000min 0.000 ng/ml

response 0

(8) Heptachlor epoxide #2 (B)

4.923min 89.636 ng/ml

response 598097909

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081224\
Data File : PD084320.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2024 20:49
Operator : AR\AJ
Sample : P3171-08RX
Misc :
ALS Vial : 43 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
DCVX0RX

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :Ankita Jodhani 08/14/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 13 10:25:47 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081224CLP.M
Quant Title : GC Extractables
QLast Update : Mon Aug 12 16:40:36 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(8) Heptachlor epoxide (B)

5.734min 22.200 ng/ml m

response 37731662

(8) Heptachlor epoxide #2 (B)

4.922min 82.321 ng/ml m

response 549291232

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081224\
Data File : PD084320.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2024 20:49
Operator : AR\AJ
Sample : P3171-08RX
Misc :
ALS Vial : 43 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
DCVX0RX

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :Ankita Jodhani 08/14/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 13 10:25:47 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081224CLP.M
Quant Title : GC Extractables
QLast Update : Mon Aug 12 16:40:36 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(10) trans-Chlordane (B)

5.983min 6.908 ng/ml

response 11405964

(10) trans-Chlordane #2 (B)

5.175min 29.761 ng/ml

response 189642101

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081224\
Data File : PD084320.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2024 20:49
Operator : AR\AJ
Sample : P3171-08RX
Misc :
ALS Vial : 43 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
DCVX0RX

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :Ankita Jodhani 08/14/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 13 10:25:47 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081224CLP.M
Quant Title : GC Extractables
QLast Update : Mon Aug 12 16:40:36 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(10) trans-Chlordane (B)
5.980min 64.925 ng/ml m
response 107200597

(10) trans-Chlordane #2 (B)
5.173min 14.494 ng/ml m
response 92361099

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081224\
Data File : PD084320.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2024 20:49
Operator : AR\AJ
Sample : P3171-08RX
Misc :
ALS Vial : 43 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
DCVX0RX

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :Ankita Jodhani 08/14/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 13 10:25:47 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081224CLP.M
Quant Title : GC Extractables
QLast Update : Mon Aug 12 16:40:36 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(13) Dieldrin (MA)

6.389min 109.304 ng/ml

response 196379524

(13) Dieldrin #2 (MA)

5.578min 42.780 ng/ml

response 263169066

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081224\
Data File : PD084320.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2024 20:49
Operator : AR\AJ
Sample : P3171-08RX
Misc :
ALS Vial : 43 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
DCVX0RX

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :Ankita Jodhani 08/14/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 13 10:25:47 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081224CLP.M
Quant Title : GC Extractables
QLast Update : Mon Aug 12 16:40:36 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(13) Dieldrin (MA)

6.389min 134.330 ng/ml m
response 241340918

(13) Dieldrin #2 (MA)

5.578min 42.780 ng/ml
response 263169066

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081224\
Data File : PD084320.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2024 20:49
Operator : AR\AJ
Sample : P3171-08RX
Misc :
ALS Vial : 43 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
DCVX0RX

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :Ankita Jodhani 08/14/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 13 10:25:47 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081224CLP.M
Quant Title : GC Extractables
QLast Update : Mon Aug 12 16:40:36 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(14) Endrin (MA)

6.624min -90.195 ng/ml

response -135115513

(14) Endrin #2 (MA)

5.850min 12.628 ng/ml

response 69715464

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081224\
Data File : PD084320.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2024 20:49
Operator : AR\AJ
Sample : P3171-08RX
Misc :
ALS Vial : 43 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :


DCVX0RX

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :Ankita Jodhani 08/14/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 13 10:25:47 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081224CLP.M
Quant Title : GC Extractables
QLast Update : Mon Aug 12 16:40:36 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(14) Endrin (MA)

6.620min 102.822 ng/ml m

response 154031409

(14) Endrin #2 (MA)

5.849min 46.372 ng/ml m

response 256012386

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081224\
Data File : PD084320.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2024 20:49
Operator : AR\AJ
Sample : P3171-08RX
Misc :
ALS Vial : 43 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

DCVX0RX


Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 08/14/2024

Supervised By :Ankita Jodhani 08/14/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 13 10:25:47 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081224CLP.M
Quant Title : GC Extractables
QLast Update : Mon Aug 12 16:40:36 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(27) Decachlorobiphenyl (SA)

9.117min 10.512 ng/ml

response 18284125

(27) Decachlorobiphenyl #2 (SA)

0.000min 0.000 ng/ml

response 0

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081224\
Data File : PD084320.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2024 20:49
Operator : AR\AJ
Sample : P3171-08RX
Misc :
ALS Vial : 43 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
DCVX0RX

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :Ankita Jodhani 08/14/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 13 10:25:47 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081224CLP.M
Quant Title : GC Extractables
QLast Update : Mon Aug 12 16:40:36 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(27) Decachlorobiphenyl (SA)

9.114min 42.688 ng/ml m

response 74247954

(27) Decachlorobiphenyl #2 (SA)

8.176min 47.000 ng/ml m

response 176623675