

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081323\
 Data File : PD077145.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Aug 2023 04:15
 Operator : AR\AJ
 Sample : 03958-07
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 NEVINS-SB02-Z10-12

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 08/14/2023
 Supervised By :Ankita Jodhani 08/14/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 14 04:36:22 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD080323.M
 Quant Title : GC Extractables
 QLast Update : Thu Aug 03 14:17:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.542	2.765	15659757	12722489	6.617m	9.340 #
28) SA Decachlor...	9.082	7.902	30539792	14225587	11.145m	9.940

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081323\
Data File : PD077145.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Aug 2023 04:15
Operator : AR\AJ
Sample : 03958-07
Misc :
ALS Vial : 33 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
NEVINS-SB02-Z10-12


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 08/14/2023
Supervised By : Ankita Jodhani 08/14/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 14 04:36:22 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD080323.M
Quant Title : GC Extractables
QLast Update : Thu Aug 03 14:17:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.542 min
Delta R.T.: 0.000 min
Response: 15659757
Conc: 6.62 ng/ml

Instrument :
ECD_D
ClientSampleId :
NEVINS-SB02-Z10-12


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/14/2023
Supervised By :Ankita Jodhani 08/14/2023




#1 Tetrachloro-m-xylene

R.T.: 2.765 min
Delta R.T.: 0.002 min
Response: 12722489
Conc: 9.34 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.082 min
Delta R.T.: -0.003 min
Response: 30539792
Conc: 11.15 ng/ml m



#28 Decachlorobiphenyl

R.T.: 7.902 min
Delta R.T.: -0.002 min
Response: 14225587
Conc: 9.94 ng/ml