

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD100918CLP.M
 Title : GC Extractables
 Last Update : Tue Oct 09 07:42:07 2018
 Response Via : Initial Calibration

Calibration Files

3 =PD049702.D 1 =PD049700.D 2 =PD049701.D
 4 =PD049703.D 5 =PD049704.D

	Compound	3	1	2	4	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.278	1.302	1.309	1.296	1.331	1.303 E6	1.49
2) A	alpha-BHC						0.000	-1.00
3) MA	gamma-BHC (Lindane)						0.000	-1.00
4) MA	Heptachlor						0.000	-1.00
5) MB	Aldrin						0.000	-1.00
6) B	beta-BHC						0.000	-1.00
7) B	delta-BHC						0.000	-1.00
8) B	Heptachlor epoxide						0.000	-1.00
9) A	Endosulfan I						0.000	-1.00
10) B	trans-Chlordane						0.000	-1.00
11) B	cis-Chlordane						0.000	-1.00
12) B	4,4'-DDE						0.000	-1.00
13) MA	Dieldrin						0.000	-1.00
14) MA	Endrin						0.000	-1.00
15) B	Endosulfan II						0.000	-1.00
16) A	4,4'-DDD						0.000	-1.00
17) MA	4,4'-DDT						0.000	-1.00
18) B	Endrin aldehyde						0.000	-1.00
19) B	Endosulfan Sulfate						0.000	-1.00
20) A	Methoxychlor						0.000	-1.00
21) B	Endrin ketone						0.000	-1.00
22)	Toxaphene-1	6.775	7.776	8.150	7.857	8.067	7.725 E3	7.15
23)	Toxaphene-2	0.946	1.023	0.955	0.966	1.014	0.981 E4	3.60
24)	Toxaphene-3	2.619	2.574	2.558	2.780	2.956	2.697 E4	6.27
25)	Toxaphene-4	2.756	2.656	2.690	2.851	3.097	2.810 E4	6.29
26)	Toxaphene-5	3.191	3.117	3.167	3.352	3.514	3.268 E4	4.99
27) SA	Decachlorobiphenyl	1.276	1.392	1.337	1.237	1.244	1.297 E6	5.09

Signal #2 Calibration Files

3 =PD049702.D 1 =PD049700.D 2 =PD049701.D
 4 =PD049703.D 5 =PD049704.D

	Compound	3	1	2	4	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.765	1.770	1.827	1.710	1.659	1.746 E7	3.67
2) A	alpha-BHC						0.000	-1.00
3) MA	gamma-BHC (Lindane)						0.000	-1.00
4) MA	Heptachlor						0.000	-1.00
5) MB	Aldrin						0.000	-1.00
6) B	beta-BHC						0.000	-1.00
7) B	delta-BHC						0.000	-1.00
8) B	Heptachlor epoxide						0.000	-1.00
9) A	Endosulfan I						0.000	-1.00
10) B	trans-Chlordane						0.000	-1.00
11) B	cis-Chlordane						0.000	-1.00
12) B	4,4'-DDE						0.000	-1.00
13) MA	Dieldrin						0.000	-1.00
14) MA	Endrin						0.000	-1.00
15) B	Endosulfan II						0.000	-1.00
16) A	4,4'-DDD						0.000	-1.00
17) MA	4,4'-DDT						0.000	-1.00
18) B	Endrin aldehyde						0.000	-1.00
19) B	Endosulfan Sulfate						0.000	-1.00
20) A	Methoxychlor						0.000	-1.00
21) B	Endrin ketone						0.000	-1.00
22)	Toxaphene-1	8.832	9.468	9.510	8.763	8.401	8.995 E4	5.34
23)	Toxaphene-2	1.057	1.116	1.126	1.034	1.002	1.067 E5	4.99

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
Method File : PD100918CLP.M
Title : GC Extractables
Last Update : Tue Oct 09 07:42:07 2018
Response Via : Initial Calibration

Calibration Files

3 =PD049702.D 1 =PD049700.D 2 =PD049701.D
4 =PD049703.D 5 =PD049704.D

	Compound	3	1	2	4	5	Avg		%RSD
24)	Toxaphene-3	1.453	1.489	1.464	1.401	1.364	1.434	E5	3.54
25)	Toxaphene-4	1.060	1.025	1.031	1.024	1.037	1.035	E5	1.42
26)	Toxaphene-5	1.818	1.830	1.843	1.710	1.783	1.797	E5	2.98
27) SA	Decachlorobiphenyl	1.593	1.718	1.673	1.505	1.466	1.591	E7	6.73

(#) = Out of Range