

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD100918CLP.M
 Title : GC Extractables
 Last Update : Tue Oct 09 08:05:00 2018
 Response Via : Initial Calibration

Calibration Files

3 =PD049709.D 1 =PD049705.D 2 =PD049707.D
 4 =PD049711.D 5 =PD049713.D

	Compound	3	1	2	4	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.360	1.333	1.499	1.360	1.371	1.385 E6	4.73
2) A	alpha-BHC	1.997	1.772	2.201	2.111	2.229	2.062 E6	9.01
3) MA	gamma-BHC (Lindane)	1.916	1.805	2.149	1.980	2.061	1.982 E6	6.67
4) MA	Heptachlor	1.978	1.877	2.227	2.036	2.103	2.044 E6	6.44
5) MB	Aldrin						0.000	-1.00
6) B	beta-BHC						0.000	-1.00
7) B	delta-BHC						0.000	-1.00
8) B	Heptachlor epoxide						0.000	-1.00
9) A	Endosulfan I	1.616	1.577	1.849	1.643	1.670	1.671 E6	6.30
10) B	trans-Chlordane						0.000	-1.00
11) B	cis-Chlordane						0.000	-1.00
12) B	4,4'-DDE						0.000	-1.00
13) MA	Dieldrin	1.787	1.623	1.973	1.883	1.947	1.843 E6	7.71
14) MA	Endrin	1.526	1.400	1.684	1.583	1.626	1.564 E6	6.91
15) B	Endosulfan II						0.000	-1.00
16) A	4,4'-DDD	1.306	1.193	1.456	1.365	1.407	1.345 E6	7.53
17) MA	4,4'-DDT	1.363	1.213	1.492	1.432	1.494	1.399 E6	8.37
18) B	Endrin aldehyde						0.000	-1.00
19) B	Endosulfan Sulfate						0.000	-1.00
20) A	Methoxychlor	7.664	7.335	8.195	7.718	7.659	7.714 E5	4.00
21) B	Endrin ketone						0.000	-1.00
22)	Toxaphene-1						0.000	-1.00
23)	Toxaphene-2						0.000	-1.00
24)	Toxaphene-3						0.000	-1.00
25)	Toxaphene-4						0.000	-1.00
26)	Toxaphene-5						0.000	-1.00
27) SA	Decachlorobiphenyl	1.395	1.482	1.564	1.309	1.246	1.399 E6	9.13

Signal #2 Calibration Files

3 =PD049709.D 1 =PD049705.D 2 =PD049707.D
 4 =PD049711.D 5 =PD049713.D

	Compound	3	1	2	4	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.837	1.852	1.665	1.802	1.727	1.777 E7	4.45
2) A	alpha-BHC	2.726	2.643	2.426	2.728	2.660	2.636 E7	4.70
3) MA	gamma-BHC (Lindane)	2.534	2.537	2.259	2.505	2.433	2.454 E7	4.76
4) MA	Heptachlor	2.543	2.589	2.269	2.490	2.385	2.455 E7	5.24
5) MB	Aldrin						0.000	-1.00
6) B	beta-BHC						0.000	-1.00
7) B	delta-BHC						0.000	-1.00
8) B	Heptachlor epoxide						0.000	-1.00
9) A	Endosulfan I	1.992	2.038	1.743	1.943	1.841	1.911 E7	6.22
10) B	trans-Chlordane						0.000	-1.00
11) B	cis-Chlordane						0.000	-1.00
12) B	4,4'-DDE						0.000	-1.00
13) MA	Dieldrin	2.088	2.125	1.864	2.054	1.967	2.019 E7	5.20
14) MA	Endrin	1.535	1.548	1.356	1.504	1.429	1.474 E7	5.46
15) B	Endosulfan II						0.000	-1.00
16) A	4,4'-DDD	1.631	1.634	1.460	1.605	1.543	1.575 E7	4.69
17) MA	4,4'-DDT	1.505	1.453	1.323	1.502	1.467	1.450 E7	5.14
18) B	Endrin aldehyde						0.000	-1.00
19) B	Endosulfan Sulfate						0.000	-1.00
20) A	Methoxychlor	6.761	6.844	6.278	6.491	5.984	6.472 E6	5.45
21) B	Endrin ketone						0.000	-1.00
22)	Toxaphene-1						0.000	-1.00
23)	Toxaphene-2						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
Method File : PD100918CLP.M
Title : GC Extractables
Last Update : Tue Oct 09 08:05:00 2018
Response Via : Initial Calibration

Calibration Files

3 =PD049709.D 1 =PD049705.D 2 =PD049707.D
4 =PD049711.D 5 =PD049713.D

	Compound	3	1	2	4	5	Avg	%RSD
24)	Toxaphene-3						0.000	-1.00
25)	Toxaphene-4						0.000	-1.00
26)	Toxaphene-5						0.000	-1.00
27) SA	Decachlorobiphenyl	1.637	1.720	1.654	1.559	1.467	1.607 E7	6.04

(#) = Out of Range