

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD100918CLP.M
 Title : GC Extractables
 Last Update : Tue Oct 09 08:27:12 2018
 Response Via : Initial Calibration

Calibration Files

3 =PD049710.D 1 =PD049706.D 2 =PD049708.D
 4 =PD049712.D 5 =PD049714.D

	Compound	3	1	2	4	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.278	1.301	1.338	1.291	1.282	1.298 E6	1.85
2) A	alpha-BHC						0.000	-1.00
3) MA	gamma-BHC (Lindane)						0.000	-1.00
4) MA	Heptachlor						0.000	-1.00
5) MB	Aldrin	1.769	1.713	1.789	1.843	1.892	1.801 E6	3.82
6) B	beta-BHC	8.193	8.387	8.602	8.034	7.864	8.216 E5	3.53
7) B	delta-BHC	1.790	1.669	1.773	1.899	1.959	1.818 E6	6.23
8) B	Heptachlor epoxide	1.666	1.669	1.707	1.697	1.706	1.689 E6	1.19
9) A	Endosulfan I						0.000	-1.00
10) B	trans-Chlordane	1.676	1.675	1.714	1.717	1.752	1.707 E6	1.89
11) B	cis-Chlordane	1.643	1.651	1.686	1.664	1.691	1.667 E6	1.26
12) B	4,4'-DDE	1.567	1.479	1.560	1.653	1.714	1.595 E6	5.69
13) MA	Dieldrin						0.000	-1.00
14) MA	Endrin						0.000	-1.00
15) B	Endosulfan II	1.495	1.454	1.518	1.532	1.541	1.508 E6	2.30
16) A	4,4'-DDD						0.000	-1.00
17) MA	4,4'-DDT						0.000	-1.00
18) B	Endrin aldehyde	1.120	1.114	1.139	1.142	1.129	1.129 E6	1.06
19) B	Endosulfan Sulfate	1.343	1.340	1.371	1.353	1.346	1.351 E6	0.93
20) A	Methoxychlor						0.000	-1.00
21) B	Endrin ketone	1.523	1.485	1.542	1.539	1.534	1.524 E6	1.53
22)	Toxaphene-1						0.000	-1.00
23)	Toxaphene-2						0.000	-1.00
24)	Toxaphene-3						0.000	-1.00
25)	Toxaphene-4						0.000	-1.00
26)	Toxaphene-5						0.000	-1.00
27) SA	Decachlorobiphenyl	1.332	1.482	1.442	1.245	1.181	1.337 E6	9.54

Signal #2 Calibration Files

3 =PD049710.D 1 =PD049706.D 2 =PD049708.D
 4 =PD049712.D 5 =PD049714.D

	Compound	3	1	2	4	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.763	1.715	1.776	1.713	1.617	1.717 E7	3.64
2) A	alpha-BHC						0.000	-1.00
3) MA	gamma-BHC (Lindane)						0.000	-1.00
4) MA	Heptachlor						0.000	-1.00
5) MB	Aldrin	2.296	2.201	2.281	2.239	2.109	2.225 E7	3.36
6) B	beta-BHC	1.064	0.967	1.045	1.026	0.974	1.015 E7	4.22
7) B	delta-BHC	2.581	2.464	2.539	2.528	2.410	2.504 E7	2.70
8) B	Heptachlor epoxide	2.114	2.059	2.112	2.033	1.896	2.043 E7	4.36
9) A	Endosulfan I						0.000	-1.00
10) B	trans-Chlordane	2.147	2.063	2.142	2.076	1.949	2.075 E7	3.87
11) B	cis-Chlordane	2.036	1.967	2.026	1.957	1.833	1.964 E7	4.12
12) B	4,4'-DDE	2.011	1.944	2.003	1.957	1.853	1.954 E7	3.24
13) MA	Dieldrin						0.000	-1.00
14) MA	Endrin						0.000	-1.00
15) B	Endosulfan II	1.734	1.685	1.742	1.667	1.561	1.678 E7	4.32
16) A	4,4'-DDD						0.000	-1.00
17) MA	4,4'-DDT						0.000	-1.00
18) B	Endrin aldehyde	1.384	1.343	1.392	1.316	1.232	1.334 E7	4.83
19) B	Endosulfan Sulfate	1.701	1.672	1.706	1.638	1.545	1.652 E7	3.97
20) A	Methoxychlor						0.000	-1.00
21) B	Endrin ketone	1.545	1.507	1.557	1.486	1.385	1.496 E7	4.58
22)	Toxaphene-1						0.000	-1.00
23)	Toxaphene-2						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
Method File : PD100918CLP.M
Title : GC Extractables
Last Update : Tue Oct 09 08:27:12 2018
Response Via : Initial Calibration

Calibration Files

3 =PD049710.D 1 =PD049706.D 2 =PD049708.D
4 =PD049712.D 5 =PD049714.D

	Compound	3	1	2	4	5	Avg	%RSD
24)	Toxaphene-3						0.000	-1.00
25)	Toxaphene-4						0.000	-1.00
26)	Toxaphene-5						0.000	-1.00
27) SA	Decachlorobiphenyl	1.598	1.668	1.655	1.498	1.379	1.560 E7	7.76

(#) = Out of Range