

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010220\
 Data File : PL055455.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jan 2020 12:38
 Operator : SG\AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 02 22:04:54 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122119.M
 Quant Title : GC Extractables
 QLast Update : Sat Dec 21 05:21:06 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds						
10) B	gamma-Chl...	0.000	6.170	0	1323649	0.499 #
17) MA	4,4'-DDT	0.000	7.155f	0	2226927	1.135 #
25)	Chlordane-3	0.000	6.170	0	1323649	4.113 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010220\
Data File : PL055455.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jan 2020 12:38
Operator : SG\AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 02 22:04:54 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122119.M
Quant Title : GC Extractables
QLast Update : Sat Dec 21 05:21:06 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 0.000 min
Exp R.T. : 3.535 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 4.047 min
Response: 0
Conc: N.D.




#10 gamma-Chlordane

R.T.: 0.000 min
Exp R.T. : 5.492 min
Response: 0
Conc: N.D.



#10 gamma-Chlordane


R.T.: 6.170 min
Delta R.T.: 0.011 min
Response: 1323649
Conc: 0.50 ng/ml



#17 4,4'-DDT


R.T.: 0.000 min
Exp R.T.: 6.473 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#17 4,4'-DDT

R.T.: 7.155 min
Delta R.T.: -0.025 min
Response: 2226927
Conc: 1.13 ng/ml




#25 Chlordane-3

R.T.: 0.000 min
Exp R.T.: 5.492 min
Response: 0
Conc: N.D.



#25 Chlordane-3

R.T.: 6.170 min
Delta R.T.: 0.011 min
Response: 1323649
Conc: 4.11 ng/ml



#28 Decachlorobiphenyl


R.T.: 0.000 min

Exp R.T. : 8.310 min

Response: 0

Conc: N.D.

Instrument :
ECD_L
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 0.000 min

Exp R.T. : 9.143 min

Response: 0

Conc: N.D.