

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010220\
 Data File : PL055483.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jan 2020 19:56
 Operator : SG\AJ
 Sample : L1001-02
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PAMT-RORO-COMP-1-12-30-19

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 02 22:10:22 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122119.M
 Quant Title : GC Extractables
 QLast Update : Sat Dec 21 05:21:06 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.534	4.048	137.9E6	39036367	13.116	14.774
28)	SA Decachlor...	8.313	9.149	88355170	22351030	9.522	10.125
Target Compounds							
6)	B beta-BHC	0.000	4.878f	0	2251465	N.D.	1.371 #
10)	B gamma-Chl...	0.000	6.169	0	1981089	N.D.	0.746 #
12)	B 4,4'-DDE	0.000	6.391	0	2806726	N.D.	1.151 #
15)	B Endosulfa...	0.000	6.982	0	2751181	N.D.	1.160 #
21)	B Endrin ke...	0.000	7.784	0	2744959	N.D.	1.233 #
25)	Chlordane-3	0.000	6.169	0	1981089	N.D.	6.156 #
27)	Chlordane-5	0.000	7.060f	0	5522689	N.D.	79.152 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010220\
Data File : PL055483.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jan 2020 19:56
Operator : SG\AJ
Sample : L1001-02
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PAMT-RORO-COMP-1-12-30-19

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 02 22:10:22 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122119.M
Quant Title : GC Extractables
QLast Update : Sat Dec 21 05:21:06 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.534 min
Delta R.T.: 0.000 min
Response: 137930236
Conc: 13.12 ng/ml

Instrument :
ECD_L
ClientSampleId :
PAMT-RORO-COMP-1-12-30-19




#1 Tetrachloro-m-xylene

R.T.: 4.048 min
Delta R.T.: 0.000 min
Response: 39036367
Conc: 14.77 ng/ml




#6 beta-BHC

R.T.: 0.000 min
Exp R.T. : 4.514 min
Response: 0
Conc: N.D.



#6 beta-BHC


R.T.: 4.878 min
Delta R.T.: -0.016 min
Response: 2251465
Conc: 1.37 ng/ml



#10 gamma-Chlordane


R.T.: 0.000 min
Exp R.T.: 5.492 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
PAMT-RORO-COMP-1-12-30-19




#10 gamma-Chlordane

R.T.: 6.169 min
Delta R.T.: 0.009 min
Response: 1981089
Conc: 0.75 ng/ml




#12 4,4'-DDE

R.T.: 0.000 min
Exp R.T.: 5.713 min
Response: 0
Conc: N.D.



#12 4,4'-DDE


R.T.: 6.391 min
Delta R.T.: 0.002 min
Response: 2806726
Conc: 1.15 ng/ml



#15 Endosulfan II


R.T.: 0.000 min
Exp R.T.: 6.381 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
PAMT-RORO-COMP-1-12-30-19




#15 Endosulfan II

R.T.: 6.982 min
Delta R.T.: 0.013 min
Response: 2751181
Conc: 1.16 ng/ml




#21 Endrin ketone

R.T.: 0.000 min
Exp R.T.: 7.256 min
Response: 0
Conc: N.D.



#21 Endrin ketone


R.T.: 7.784 min
Delta R.T.: -0.006 min
Response: 2744959
Conc: 1.23 ng/ml



#25 Chlordane-3


R.T.: 0.000 min
Exp R.T.: 5.492 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
PAMT-RORO-COMP-1-12-30-19




#25 Chlordane-3

R.T.: 6.169 min
Delta R.T.: 0.009 min
Response: 1981089
Conc: 6.16 ng/ml




#27 Chlordane-5

R.T.: 0.000 min
Exp R.T.: 6.391 min
Response: 0
Conc: N.D.



#27 Chlordane-5


R.T.: 7.060 min
Delta R.T.: 0.019 min
Response: 5522689
Conc: 79.15 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.313 min
Delta R.T.: 0.002 min
Response: 88355170
Conc: 9.52 ng/ml

Instrument :
ECD_L
ClientSampleId :
PAMT-RORO-COMP-1-12-30-19



#28 Decachlorobiphenyl

R.T.: 9.149 min
Delta R.T.: 0.006 min
Response: 22351030
Conc: 10.13 ng/ml