

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010220\
 Data File : PL055484.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jan 2020 20:10
 Operator : SG\AJ
 Sample : L1001-03
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 02 22:10:32 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122119.M
 Quant Title : GC Extractables
 QLast Update : Sat Dec 21 05:21:06 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.534	4.048	154.0E6	43600630	14.648	16.501
28)	SA Decachlor...	8.314	9.149	102.1E6	24949374	11.008	11.302
Target Compounds							
12)	B 4,4'-DDE	0.000	6.392	0	1314487	N.D.	0.539 #
15)	B Endosulfa...	0.000	6.982	0	1071642	N.D.	0.452 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010220\
Data File : PL055484.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jan 2020 20:10
Operator : SG\AJ
Sample : L1001-03
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 02 22:10:32 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122119.M
Quant Title : GC Extractables
QLast Update : Sat Dec 21 05:21:06 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.534 min
Delta R.T.: 0.000 min
Response: 154041674
Conc: 14.65 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.048 min
Delta R.T.: 0.000 min
Response: 43600630
Conc: 16.50 ng/ml




#12 4,4'-DDE

R.T.: 0.000 min
Exp R.T. : 5.713 min
Response: 0
Conc: N.D.



#12 4,4'-DDE


R.T.: 6.392 min
Delta R.T.: 0.003 min
Response: 1314487
Conc: 0.54 ng/ml



#15 Endosulfan II


R.T.: 0.000 min
Exp R.T.: 6.381 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#15 Endosulfan II

R.T.: 6.982 min
Delta R.T.: 0.014 min
Response: 1071642
Conc: 0.45 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.314 min
Delta R.T.: 0.003 min
Response: 102143423
Conc: 11.01 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.149 min
Delta R.T.: 0.006 min
Response: 24949374
Conc: 11.30 ng/ml