

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010220\
 Data File : PL055478.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jan 2020 18:46
 Operator : SG\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 100 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 02 22:09:13 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122119.M
 Quant Title : GC Extractables
 QLast Update : Sat Dec 21 05:21:06 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.535	4.048	489.0E6	136.8E6	46.501	51.766
28)	SA Decachlor...	8.312	9.149	439.1E6	106.2E6	47.318	48.127
Target Compounds							
2)	A alpha-BHC	3.978	4.440	812.3E6	198.7E6	46.693	53.964
3)	MA gamma-BHC...	4.264	4.725	734.1E6	181.4E6	44.520	53.070
4)	MA Heptachlor	4.569	5.234	764.2E6	165.1E6	46.807	52.482
5)	MB Aldrin	4.821	5.540	719.6E6	153.5E6	47.126	53.302
6)	B beta-BHC	4.515	4.896	317.8E6	81580831	46.339	49.678
7)	B delta-BHC	4.723	5.114	725.5E6	152.7E6	47.930	54.619
8)	B Heptachlo...	5.265	5.927	659.2E6	133.1E6	46.135	52.231
9)	A Endosulfan I	5.607	6.285	613.1E6	125.9E6	46.402	50.838
10)	B gamma-Chl...	5.493	6.162	669.5E6	137.3E6	45.951	51.730
11)	B alpha-Chl...	5.552	6.234	647.9E6	134.1E6	46.022	51.136
12)	B 4,4'-DDE	5.715	6.391	629.8E6	126.4E6	47.262	51.833
13)	MA Dieldrin	5.851	6.543	660.9E6	132.5E6	47.340	51.906
14)	MA Endrin	6.110	6.762	571.4E6	110.0E6	44.958	51.246
15)	B Endosulfa...	6.383	6.972	524.5E6	119.2E6	44.692	50.277
16)	A 4,4'-DDD	6.232	6.884	465.6E6	101.0E6	48.477	53.147
17)	MA 4,4'-DDT	6.474	7.183	486.0E6	100.2E6	48.275	51.055
18)	B Endrin al...	6.551	7.096	471.7E6	99722814	48.436	51.672
19)	B Endosulfa...	6.765	7.321	468.7E6	105.8E6	46.595	50.680
20)	A Methoxychlor	7.019	7.644	224.4E6	54522987	46.564	50.518
21)	B Endrin ke...	7.258	7.794	521.6E6	114.8E6	49.762	51.531
22)	Mirex	7.450	8.251	345.0E6	80923872	45.296	48.806

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010220\
Data File : PL055478.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jan 2020 18:46
Operator : SG\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 100 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 02 22:09:13 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122119.M
Quant Title : GC Extractables
QLast Update : Sat Dec 21 05:21:06 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

