

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010320\
 Data File : PL055497.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jan 2020 18:30
 Operator : AJ\MA
 Sample : PSTDICC075
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC075

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 04 00:14:35 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL010320.M
 Quant Title : GC Extractables
 QLast Update : Sat Jan 04 00:13:56 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.501	4.012	742.0E6	210.0E6	79.322	76.250
2)	SA Decachlor...	8.266	9.083	637.3E6	161.1E6	75.654	74.132
Target Compounds							
2)	A alpha-BHC	3.942	4.402	1261.8E6	326.6E6	81.892	77.427
3)	MA gamma-BHC...	4.226	4.686	1182.4E6	298.8E6	80.743	77.006
4)	MA Heptachlor	4.530	5.192	1147.2E6	263.7E6	79.527	75.720
5)	MB Aldrin	4.779	5.495	1094.5E6	240.8E6	79.054	75.591
6)	B beta-BHC	4.477	4.856	461.5E6	121.6E6	79.377	74.381
7)	B delta-BHC	4.683	5.072	1122.6E6	233.2E6	80.550	81.434
8)	B Heptachlo...	5.223	5.881	985.8E6	189.5E6	79.849	77.633
9)	A Endosulfan I	5.563	6.237	921.4E6	199.9E6	78.220	75.235
10)	B gamma-Chl...	5.450	6.115	995.3E6	210.6E6	78.690	75.475
11)	B alpha-Chl...	5.508	6.188	968.0E6	211.3E6	76.888	75.675
12)	B 4,4'-DDE	5.672	6.345	901.6E6	203.3E6	78.757	76.077
13)	MA Dieldrin	5.806	6.495	984.0E6	208.3E6	79.077	76.032
14)	MA Endrin	6.064	6.713	897.7E6	179.5E6	78.875	75.791
15)	B Endosulfa...	6.337	6.923	815.6E6	182.7E6	109.869	74.502 #
16)	A 4,4'-DDD	6.188	6.836	659.6E6	159.5E6	78.923	75.701
17)	MA 4,4'-DDT	6.429	7.135	727.6E6	168.2E6	79.935	74.832
18)	B Endrin al...	6.505	7.048	676.2E6	150.2E6	78.622	75.574
19)	B Endosulfa...	6.718	7.271	738.7E6	164.5E6	111.675	74.646 #
20)	A Methoxychlor	6.974	7.596	356.1E6	87864162	78.241	74.501
21)	B Endrin ke...	7.211	7.744	836.5E6	178.0E6	78.968	75.504
22)	Mirex	7.400	8.199	535.4E6	119.3E6	121.894	73.134 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010320\
Data File : PL055497.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jan 2020 18:30
Operator : AJ\MA
Sample : PSTDICC075
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC075

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 04 00:14:35 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL010320.M
Quant Title : GC Extractables
QLast Update : Sat Jan 04 00:13:56 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

