

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010323\
 Data File : PL080043.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jan 2023 18:20
 Operator : AR\AJ
 Sample : N6242-01MS
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 MM-1MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 03 20:37:08 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL123022.M
 Quant Title : GC Extractables
 QLast Update : Fri Dec 30 05:17:19 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.366	2.695	33546740	48123381	24.183	26.191
28)	SA Decachlor...	8.817	7.852	38162681	49902483	27.145	29.843
Target Compounds							
2)	A alpha-BHC	3.815	3.197	109.4E6	160.0E6	57.104	63.743
3)	MA gamma-BHC...	4.144	3.526	101.8E6	150.8E6	58.396	63.146
4)	MA Heptachlor	4.729	3.869	107.8E6	152.8E6	60.033	71.073
5)	MB Aldrin	5.068	4.149	106.2E6	157.2E6	63.539	68.602
6)	B beta-BHC	4.340	3.821	47257220	64391708	55.795	60.019
7)	B delta-BHC	4.583	4.051	88751879	128.5E6	52.406	56.185
8)	B Heptachlo...	5.494	4.652	101.6E6	138.6E6	61.791	65.564
9)	A Endosulfan I	5.876	5.021	98242987	129.3E6	63.026	63.141
10)	B gamma-Chl...	5.748	4.903	106.2E6	142.6E6	61.947	66.271
11)	B alpha-Chl...	5.828	4.967	105.9E6	142.2E6	63.732	66.584
12)	B 4,4'-DDE	6.006	5.160	95927316	125.4E6	65.018	68.854
13)	MA Dieldrin	6.151	5.287	106.6E6	139.2E6	65.236	68.950
14)	MA Endrin	6.378	5.561	91341057	118.8E6	69.641	76.728
15)	B Endosulfa...	6.597	5.855	93833393	119.1E6	64.950	68.610
16)	A 4,4'-DDD	6.520	5.714	81843129	98187169	65.275	64.447
17)	MA 4,4'-DDT	6.834	5.966	85876231	104.4E6	67.702	79.493
18)	B Endrin al...	6.727	6.035	77554065	98053372	70.946	72.874
19)	B Endosulfa...	6.962	6.258	91628901	114.6E6	63.298	65.789
20)	A Methoxychlor	7.313	6.546	50379018	60206202	71.080	78.182
21)	B Endrin ke...	7.442	6.763	99681995	125.2E6	65.742	67.803
22)	Mirex	7.914	6.949	83825491	109.8E6	63.765	71.319

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010323\
Data File : PL080043.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jan 2023 18:20
Operator : AR\AJ
Sample : N6242-01MS
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
MM-1MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 03 20:37:08 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL123022.M
Quant Title : GC Extractables
QLast Update : Fri Dec 30 05:17:19 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

