

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010323\
 Data File : PL080036.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jan 2023 16:43
 Operator : AR\AJ
 Sample : N6235-01
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 OR-2-122922

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 01/04/2023
 Supervised By : Ankita Jodhani 01/04/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 03 20:33:17 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL123022.M
 Quant Title : GC Extractables
 QLast Update : Fri Dec 30 05:17:19 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.366	2.695	17244543	25304904	12.431	13.772
28)	SA Decachlor...	8.819	7.853	22165552	26802011	15.766	16.028
Target Compounds							
17)	MA 4,4'-DDT	6.834	5.965	729819	1223482	0.575m	0.931 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010323\
Data File : PL080036.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jan 2023 16:43
Operator : AR\AJ
Sample : N6235-01
Misc :
ALS Vial : 16 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
OR-2-122922


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/04/2023
Supervised By :Ankita Jodhani 01/04/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 03 20:33:17 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL123022.M
Quant Title : GC Extractables
QLast Update : Fri Dec 30 05:17:19 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.366 min
Delta R.T.: 0.001 min
Response: 17244543
Conc: 12.43 ng/ml

Instrument :
ECD_L
ClientSampleId :
OR-2-122922


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/04/2023
Supervised By :Ankita Jodhani 01/04/2023




#1 Tetrachloro-m-xylene

R.T.: 2.695 min
Delta R.T.: 0.000 min
Response: 25304904
Conc: 13.77 ng/ml




#17 4,4'-DDT

R.T.: 6.834 min
Delta R.T.: 0.006 min
Response: 729819
Conc: 0.58 ng/ml m



#17 4,4'-DDT

R.T.: 5.965 min
Delta R.T.: -0.004 min
Response: 1223482
Conc: 0.93 ng/ml




#28 Decachlorobiphenyl

R.T.: 8.819 min
Delta R.T.: 0.011 min
Response: 22165552
Conc: 15.77 ng/ml

Instrument :
ECD_L
ClientSampleId :
OR-2-122922

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/04/2023
Supervised By :Ankita Jodhani 01/04/2023



#28 Decachlorobiphenyl

R.T.: 7.853 min
Delta R.T.: -0.003 min
Response: 26802011
Conc: 16.03 ng/ml