

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010323\
 Data File : PL080037.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jan 2023 16:57
 Operator : AR\AJ
 Sample : N6237-01
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TR-05-122922

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 03 20:33:42 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL123022.M
 Quant Title : GC Extractables
 QLast Update : Fri Dec 30 05:17:19 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.366	2.695	17593818	25244528	12.683	13.739
28)	SA Decachlor...	8.819	7.853	21099367	26864591	15.008	16.066
Target Compounds							
12)	B 4,4'-DDE	6.007	5.161	16341675	18774617	11.076	10.307
17)	MA 4,4'-DDT	6.835	5.963	1650157	6248824	1.301	4.757 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010323\
Data File : PL080037.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jan 2023 16:57
Operator : AR\AJ
Sample : N6237-01
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TR-05-122922

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 03 20:33:42 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL123022.M
Quant Title : GC Extractables
QLast Update : Fri Dec 30 05:17:19 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.366 min
Delta R.T.: 0.001 min
Response: 17593818
Conc: 12.68 ng/ml

Instrument :
ECD_L
ClientSampleId :
TR-05-122922




#1 Tetrachloro-m-xylene

R.T.: 2.695 min
Delta R.T.: 0.000 min
Response: 25244528
Conc: 13.74 ng/ml




#12 4,4'-DDE

R.T.: 6.007 min
Delta R.T.: 0.006 min
Response: 16341675
Conc: 11.08 ng/ml



#12 4,4'-DDE


R.T.: 5.161 min
Delta R.T.: -0.002 min
Response: 18774617
Conc: 10.31 ng/ml



#17 4,4'-DDT


R.T.: 6.835 min
Delta R.T.: 0.007 min
Response: 1650157
Conc: 1.30 ng/ml

Instrument :
ECD_L
ClientSampleId :
TR-05-122922




#17 4,4'-DDT

R.T.: 5.963 min
Delta R.T.: -0.006 min
Response: 6248824
Conc: 4.76 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.819 min
Delta R.T.: 0.011 min
Response: 21099367
Conc: 15.01 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.853 min
Delta R.T.: -0.003 min
Response: 26864591
Conc: 16.07 ng/ml