

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010323\
 Data File : PL080041.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jan 2023 17:52
 Operator : AR\AJ
 Sample : N6239-04
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TP-C1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 03 20:35:58 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL123022.M
 Quant Title : GC Extractables
 QLast Update : Fri Dec 30 05:17:19 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.366	2.695	18186244	26747493	13.110	14.557
28)	SA Decachlor...	8.817	7.853	23595389	29921847	16.783	17.894

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010323\
Data File : PL080041.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jan 2023 17:52
Operator : AR\AJ
Sample : N6239-04
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TP-C1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 03 20:35:58 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL123022.M
Quant Title : GC Extractables
QLast Update : Fri Dec 30 05:17:19 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.366 min
Delta R.T.: 0.000 min
Response: 18186244
Conc: 13.11 ng/ml

Instrument :
ECD_L
ClientSampleId :
TP-C1




#1 Tetrachloro-m-xylene

R.T.: 2.695 min
Delta R.T.: -0.001 min
Response: 26747493
Conc: 14.56 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.817 min
Delta R.T.: 0.009 min
Response: 23595389
Conc: 16.78 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.853 min
Delta R.T.: -0.003 min
Response: 29921847
Conc: 17.89 ng/ml