

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010323\
 Data File : PL080049.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jan 2023 20:11
 Operator : AR\AJ
 Sample : N6242-03
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 MM-3

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 01/04/2023
 Supervised By : Ankita Jodhani 01/04/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 03 23:24:05 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL123022.M
 Quant Title : GC Extractables
 QLast Update : Fri Dec 30 05:17:19 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.365	2.693	39052830	56773638	28.152	30.898m
28) SA Decachlor...	8.817	7.852	42531508	55629019	30.253	33.268

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010323\
Data File : PL080049.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jan 2023 20:11
Operator : AR\AJ
Sample : N6242-03
Misc :
ALS Vial : 26 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
MM-3


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 01/04/2023
Supervised By : Ankita Jodhani 01/04/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 03 23:24:05 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL123022.M
Quant Title : GC Extractables
QLast Update : Fri Dec 30 05:17:19 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.365 min
Delta R.T.: 0.000 min
Response: 39052830
Conc: 28.15 ng/ml

Instrument :
ECD_L
ClientSampleId :
MM-3


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/04/2023
Supervised By :Ankita Jodhani 01/04/2023




#1 Tetrachloro-m-xylene

R.T.: 2.693 min
Delta R.T.: -0.003 min
Response: 56773638
Conc: 30.90 ng/ml m



#28 Decachlorobiphenyl

R.T.: 8.817 min
Delta R.T.: 0.009 min
Response: 42531508
Conc: 30.25 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.852 min
Delta R.T.: -0.004 min
Response: 55629019
Conc: 33.27 ng/ml