

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010324\
 Data File : PL087622.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jan 2024 15:53
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 03 21:43:25 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122823.M
 Quant Title : GC Extractables
 QLast Update : Fri Dec 29 03:17:47 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.421	2.675	156.0E6	63948224	52.109	52.791
2)	SA Decachlor...	8.983	7.863	145.2E6	63827130	50.966	50.384
Target Compounds							
2)	A alpha-BHC	3.877	3.179	240.5E6	97721361	53.610	53.021
3)	MA gamma-BHC...	4.211	3.509	232.5E6	95602470	53.313	52.608
4)	MA Heptachlor	4.804	3.851	226.7E6	90511470	51.595	50.821
5)	MB Aldrin	5.146	4.132	226.5E6	92037444	52.853	52.149
6)	B beta-BHC	4.411	3.811	98249637	40269431	52.809	52.651
7)	B delta-BHC	4.657	4.040	229.9E6	95052737	54.639	54.756
8)	B Heptachlo...	5.578	4.639	201.5E6	82544874	52.605	51.792
9)	A Endosulfan I	5.965	5.010	184.7E6	76282217	51.225	49.311
10)	B gamma-Chl...	5.836	4.893	203.7E6	84020454	52.689	52.055
11)	B alpha-Chl...	5.917	4.957	201.2E6	83686615	52.350	50.425
12)	B 4,4'-DDE	6.098	5.155	174.2E6	73957717	54.498	54.982
13)	MA Dieldrin	6.243	5.278	194.4E6	82940420	52.186	52.192
14)	MA Endrin	6.473	5.553	151.1E6	69965702	48.103	49.342
15)	B Endosulfa...	6.695	5.851	163.7E6	71698710	51.058	52.374
16)	A 4,4'-DDD	6.619	5.714	143.8E6	64216721	55.776	56.952
17)	MA 4,4'-DDT	6.936	5.966	139.6E6	63140875	48.824	50.733
18)	B Endrin al...	6.826	6.033	124.6E6	56577024	49.536	48.680
19)	B Endosulfa...	7.063	6.258	156.6E6	69627342	51.193	53.042
20)	A Methoxychlor	7.421	6.552	74610323	35617329	45.954	48.388
21)	B Endrin ke...	7.549	6.764	173.1E6	84287020	53.206	54.980
22)	Mirex	8.028	6.945	134.9E6	65424774	48.127	48.775

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010324\
Data File : PL087622.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jan 2024 15:53
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 03 21:43:25 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122823.M
Quant Title : GC Extractables
QLast Update : Fri Dec 29 03:17:47 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

