

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\
 Method File : LTX010422.M
 Title : GC Extractables
 Last Update : Wed Jan 05 02:19:49 2022
 Response Via : Initial Calibration

Calibration Files

500 =PL071799.D 1000 =PL071797.D 250 =PL071800.D
 750 =PL071798.D 100 =PL071801.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	4.672	4.633	4.677	4.724	4.501	4.641	E5	1.83
2)	Toxaphene-1	2.789	2.687	2.802	2.814	2.977	2.814	E3	3.70
3)	Toxaphene-2	4.255	4.122	4.327	4.323	4.269	4.259	E3	1.95
4)	Toxaphene-3	1.060	1.078	1.057	1.082	1.059	1.067	E4	1.12
5)	Toxaphene-4	9.029	8.954	8.739	8.829	8.477	8.806	E3	2.45
6)	Toxaphene-5	1.403	1.425	1.423	1.459	1.372	1.416	E4	2.25
7) SA	Decachlorobiphenyl	7.105	6.566	7.558	6.941	7.784	7.191	E5	6.77

Signal #2 Calibration Files

500 =PL071799.D 1000 =PL071797.D 250 =PL071800.D
 750 =PL071798.D 100 =PL071801.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	1.259	1.259	1.255	1.282	1.199	1.251	E6	2.47
2)	Toxaphene-1	7.766	7.714	8.201	7.901	7.131	7.743	E3	5.05
3)	Toxaphene-2	1.364	1.406	1.399	1.392	1.493	1.411	E4	3.44
4)	Toxaphene-3	1.874	1.929	1.893	1.968	1.853	1.903	E4	2.39
5)	Toxaphene-4	1.780	1.833	1.792	1.840	1.614	1.772	E4	5.18
6)	Toxaphene-5	1.472	1.445	1.425	1.541	1.353	1.447	E4	4.74
7) SA	Decachlorobiphenyl	1.204	1.143	1.235	1.190	1.208	1.196	E6	2.86

(#) = Out of Range