

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010422\
 Data File : PL071787.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Jan 2022 17:14
 Operator : AR\AJ
 Sample : PSTDICC100
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 05 01:41:30 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL010422.M
 Quant Title : GC Extractables
 QLast Update : Wed Jan 05 01:41:20 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.557	5.493	45473845	122.0E6	98.754	98.062
2)	SA Decachlor...	10.315	11.382	64797937	111.3E6	94.720	95.539
Target Compounds							
2)	A alpha-BHC	5.263	6.074	67840126	186.1E6	102.116	100.604
3)	MA gamma-BHC...	5.685	6.470	64542421	169.4E6	100.755	100.004
4)	MA Heptachlor	6.091	7.116	70225148	165.7E6	99.801	99.457
5)	MB Aldrin	6.415	7.486	65367718	158.5E6	100.037	99.486
6)	B beta-BHC	6.069	6.729	25669496	68932388	95.771	96.912
7)	B delta-BHC	6.328	6.995	65816986	161.3E6	100.232	99.440
8)	B Heptachlo...	6.992	7.953	61313365	139.3E6	98.716	98.425
9)	A Endosulfan I	7.394	8.358	56802281	123.5E6	98.584	98.400
10)	B gamma-Chl...	7.269	8.223	62139275	136.5E6	98.876	98.602
11)	B alpha-Chl...	7.339	8.303	61772043	132.8E6	98.319	98.293
12)	B 4,4'-DDE	7.554	8.493	57365670	125.8E6	99.821	98.240
13)	MA Dieldrin	7.680	8.647	61988542	133.2E6	99.745	98.941
14)	MA Endrin	7.971	8.886	54064729	105.1E6	99.453	98.779
15)	B Endosulfa...	8.283	9.121	56525815	110.4E6	98.000	97.714
16)	A 4,4'-DDD	8.139	9.033	49500896	96653766	99.541	98.362
17)	MA 4,4'-DDT	8.399	9.349	50144202	97548867	99.980	99.187
18)	B Endrin al...	8.472	9.256	43400752	82554996	96.909	96.851
19)	B Endosulfa...	8.703	9.494	55688889	101.3E6	96.927	97.249
20)	A Methoxychlor	8.995	9.837	29932677	49594917	96.695	97.348
21)	B Endrin ke...	9.220	9.990	67909287	111.9E6	97.332	97.466
22)	Mirex	9.402	10.453	53535442	81361984	94.350	95.524

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010422\
Data File : PL071787.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Jan 2022 17:14
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 05 01:41:30 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL010422.M
Quant Title : GC Extractables
QLast Update : Wed Jan 05 01:41:20 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

