

Table 2. Surrogate QC Acceptance Criteria

Parameter	% Recovery	
	Water	Soil
Tetrachloro-m-xylene	30 – 150	30 - 150
Decachlorobiphenyl	30 – 150	30 - 150

Table 3. Retention Time Criteria for Single Component Pesticides

Compound	Retention Time
Aldrin	±0.05
alpha-BHC	±0.05
beta-BHC	±0.05
delta-BHC	±0.05
gamma-BHC	±0.05
Heptachlor	±0.07
Heptachlor Epoxide	±0.07
DDD	±0.07
DDE	±0.07
DDT	±0.07
Dieldrin	±0.07
Endrin	±0.07
Endrin aldehyde	±0.07
Endrin ketone	±0.07
Endosulfan I	±0.07
Endosulfan II	±0.07
Endosulfan sulfate	±0.07
Methoxychlor	±0.07
TCMX	±0.05
DCB	±0.10

Table 4. Laboratory Control Sample Recovery Limits

Compound	Percent Recovery
----------	------------------