

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010522\
 Data File : PL071817.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jan 2022 14:03
 Operator : AR\AJ
 Sample : PB141850BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB141850BL

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 01/06/2022
 Supervised By : Ankita Jodhani 01/06/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 06 00:39:09 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL010422.M
 Quant Title : GC Extractables
 QLast Update : Wed Jan 05 01:51:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.558	5.494	9258951	24961498	20.043m	20.316
2) SA Decachlor...	10.320	11.388	15053298	25076941	20.820m	21.162

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010522\
Data File : PL071817.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jan 2022 14:03
Operator : AR\AJ
Sample : PB141850BL
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
PB141850BL


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 01/06/2022
Supervised By : Ankita Jodhani 01/06/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 06 00:39:09 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL010422.M
Quant Title : GC Extractables
QLast Update : Wed Jan 05 01:51:06 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 4.558 min
Delta R.T.: 0.002 min
Response: 9258951
Conc: 20.04 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB141850BL


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/06/2022
Supervised By :Ankita Jodhani 01/06/2022




#1 Tetrachloro-m-xylene

R.T.: 5.494 min
Delta R.T.: 0.001 min
Response: 24961498
Conc: 20.32 ng/ml



#28 Decachlorobiphenyl

R.T.: 10.320 min
Delta R.T.: 0.005 min
Response: 15053298
Conc: 20.82 ng/ml m



#28 Decachlorobiphenyl

R.T.: 11.388 min
Delta R.T.: 0.007 min
Response: 25076941
Conc: 21.16 ng/ml