

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010720\
 Data File : PL055576.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Jan 2020 15:49
 Operator : AJ\MA
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

mohammad
 1/7/2020 4:16:01 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 07 16:00:56 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL010320.M
 Quant Title : GC Extractables
 QLast Update : Sat Jan 04 01:07:28 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.500	4.013	515.8E6	146.2E6	53.548	51.710
28)	SA Decachlor...	8.261	9.088	420.6E6	101.0E6	49.729	45.625
Target Compounds							
2)	A alpha-BHC	3.940	4.403	877.9E6	222.8E6	54.411	52.650
3)	MA gamma-BHC...	4.224	4.686	831.7E6	203.4E6	55.385	51.553
4)	MA Heptachlor	4.527	5.192	774.1E6	171.2E6	52.516	47.540
5)	MB Aldrin	4.776	5.496	733.9E6	157.9E6	50.687	49.042
6)	B beta-BHC	4.474	4.857	299.7E6	83771107	48.374	47.887
7)	B delta-BHC	4.680	5.073	753.6E6	129.6E6	53.415	44.516
8)	B Heptachlo...	5.219	5.882	633.8E6	113.5E6	49.677	45.877
9)	A Endosulfan I	5.559	6.240	592.1E6	131.2E6	49.223	47.631
10)	B gamma-Chl...	5.446	6.117	636.1E6	133.7E6	49.592	46.759
11)	B alpha-Chl...	5.505	6.190	624.3E6	138.6E6	49.271	47.883
12)	B 4,4'-DDE	5.668	6.347	634.4E6	133.9E6	54.650	49.141
13)	MA Dieldrin	5.802	6.498	635.2E6	133.9E6	50.324	47.867
14)	MA Endrin	6.060	6.716	578.7E6	117.3E6	49.811	48.683
15)	B Endosulfa...	6.332	6.926	524.7E6	118.8E6	52.709	45.281
16)	A 4,4'-DDD	6.184	6.839	453.7E6	109.5E6	54.520	50.992
17)	MA 4,4'-DDT	6.424	7.139	495.0E6	109.1E6	57.106	45.155
18)	B Endrin al...	6.500	7.049	430.0E6	97340126	47.924	47.027m
19)	B Endosulfa...	6.712	7.275	469.9E6	105.5E6	54.511m	46.288
20)	A Methoxychlor	6.970	7.600	243.0E6	57350912	52.708	47.594
21)	B Endrin ke...	7.206	7.747	506.4E6	111.6E6	43.698	46.143
22)	Mirex	7.396	8.203	333.6E6	77379900	53.208	45.436

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010720\
 Data File : PL055576.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Jan 2020 15:49
 Operator : AJ\MA
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

mohammad
 1/7/2020 4:16:01 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 07 16:00:56 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL010320.M
 Quant Title : GC Extractables
 QLast Update : Sat Jan 04 01:07:28 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

