

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010820\
 Data File : PL055588.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jan 2020 12:12
 Operator : AJ\MA
 Sample : PSTDICV050
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 ICVPL010820

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 08 12:23:51 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL010820.M
 Quant Title : GC Extractables
 QLast Update : Wed Jan 08 12:19:19 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.500	4.013	550.8E6	151.5E6	49.386	49.275
2)	SA Decachlor...	8.262	9.088	479.8E6	115.2E6	49.456	49.813
Target Compounds							
2)	A alpha-BHC	3.940	4.403	947.2E6	234.1E6	50.047	50.517
3)	MA gamma-BHC...	4.225	4.687	885.4E6	216.7E6	50.462	50.415
4)	MA Heptachlor	4.528	5.193	900.7E6	194.0E6	49.951	49.808
5)	MB Aldrin	4.777	5.497	860.0E6	178.6E6	50.169	49.077
6)	B beta-BHC	4.475	4.857	352.4E6	91257334	49.831	51.610
7)	B delta-BHC	4.680	5.074	860.1E6	174.9E6	50.376	51.197
8)	B Heptachlo...	5.220	5.883	792.4E6	146.1E6	49.515	47.062
9)	A Endosulfan I	5.560	6.240	741.7E6	151.4E6	48.899	49.049
10)	B gamma-Chl...	5.447	6.118	809.3E6	156.4E6	49.245	47.683
11)	B alpha-Chl...	5.505	6.190	793.6E6	160.0E6	49.070	48.861
12)	B 4,4'-DDE	5.669	6.348	743.8E6	152.8E6	49.355	49.518
13)	MA Dieldrin	5.803	6.498	792.7E6	157.2E6	49.456	49.622
14)	MA Endrin	6.061	6.717	716.2E6	132.7E6	49.091	49.741
15)	B Endosulfa...	6.334	6.926	609.0E6	132.4E6	49.521	49.601
16)	A 4,4'-DDD	6.185	6.839	541.5E6	117.6E6	48.894	50.197
17)	MA 4,4'-DDT	6.426	7.139	559.2E6	119.4E6	48.550	50.202
18)	B Endrin al...	6.502	7.051	525.3E6	112.9E6	48.233	49.513
19)	B Endosulfa...	6.715	7.275	556.3E6	123.3E6	49.654	48.892
20)	A Methoxychlor	6.971	7.599	273.6E6	63584529	49.780	49.457
21)	B Endrin ke...	7.207	7.747	617.0E6	130.5E6	50.437	49.589
22)	Mirex	7.397	8.203	401.0E6	91128810	50.058	48.575

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010820\
Data File : PL055588.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jan 2020 12:12
Operator : AJ\MA
Sample : PSTDICV050
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
ICVPL010820

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 08 12:23:51 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL010820.M
Quant Title : GC Extractables
QLast Update : Wed Jan 08 12:19:19 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

