

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010919\
 Data File : PL043699.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jan 2019 18:14
 Operator : AJ\SJ
 Sample : PSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 09 05:46:34 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL010919.M
 Quant Title : GC Extractables
 QLast Update : Wed Jan 09 05:45:47 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.407	4.035	329.3E6	147.5E6	50.000	50.000
28)	SA Decachlor...	8.119	9.133	364.7E6	130.8E6	50.000	50.000
Target Compounds							
2)	A alpha-BHC	3.842	4.427	511.9E6	221.7E6	50.000	50.000
3)	MA gamma-BHC...	4.122	4.713	462.5E6	210.6E6	50.000	50.000
4)	MA Heptachlor	4.413	5.225	454.9E6	181.7E6	50.000	50.000
5)	MB Aldrin	4.657	5.532	441.6E6	190.0E6	50.000	50.000
6)	B beta-BHC	4.376	4.884	194.2E6	88176137	50.000	50.000
7)	B delta-BHC	4.575	5.102	445.4E6	201.9E6	50.000	50.000
8)	B Heptachlo...	5.098	5.920	403.3E6	184.8E6	50.000	50.000
9)	A Endosulfan I	5.433	6.278	374.9E6	153.6E6	50.000	50.000
10)	B gamma-Chl...	5.323	6.154	411.0E6	160.0E6	50.000	50.000
11)	B alpha-Chl...	5.379	6.227	399.0E6	162.0E6	50.000	50.000
12)	B 4,4'-DDE	5.546	6.384	385.0E6	171.3E6	50.000	50.000
13)	MA Dieldrin	5.674	6.536	408.2E6	170.0E6	50.000	50.000
14)	MA Endrin	5.929	6.754	363.8E6	139.5E6	50.000	50.000
15)	B Endosulfa...	6.204	6.964	354.0E6	145.4E6	50.000	50.000
16)	A 4,4'-DDD	6.061	6.875	316.3E6	141.1E6	50.000	50.000
17)	MA 4,4'-DDT	6.298	7.175	313.7E6	143.4E6	50.000	50.000
18)	B Endrin al...	6.372	7.088	279.2E6	119.9E6	50.000	50.000
19)	B Endosulfa...	6.583	7.311	341.9E6	145.7E6	50.000	50.000
20)	A Methoxychlor	6.844	7.633	179.9E6	71007594	50.000	50.000
21)	B Endrin ke...	7.070	7.784	399.0E6	152.0E6	50.000	50.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010919\
Data File : PL043699.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jan 2019 18:14
Operator : AJ\SJ
Sample : PSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 09 05:46:34 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL010919.M
Quant Title : GC Extractables
QLast Update : Wed Jan 09 05:45:47 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

