

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010923\
 Data File : PL080194.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jan 2023 21:46
 Operator : AR\AJ
 Sample : 01066-01
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 OILY-WATER-1

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 01/10/2023
 Supervised By : Ankita Jodhani 01/10/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 09 23:36:57 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL010423.M
 Quant Title : GC Extractables
 QLast Update : Thu Jan 05 05:11:24 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.371	2.692	38391519	38168381	19.300	16.786
28) SA Decachlor...	8.885	7.844	29651141	34853424	17.282m	15.427

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL010923\
Data File : PL080194.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jan 2023 21:46
Operator : AR\AJ
Sample : 01066-01
Misc :
ALS Vial : 27 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
OILY-WATER-1


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 01/10/2023
Supervised By : Ankita Jodhani 01/10/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 09 23:36:57 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL010423.M
Quant Title : GC Extractables
QLast Update : Thu Jan 05 05:11:24 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.371 min
Delta R.T.: 0.000 min
Response: 38391519
Conc: 19.30 ng/ml

Instrument :
ECD_L
ClientSampleId :
OILY-WATER-1


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/10/2023
Supervised By :Ankita Jodhani 01/10/2023




#1 Tetrachloro-m-xylene

R.T.: 2.692 min
Delta R.T.: -0.002 min
Response: 38168381
Conc: 16.79 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.885 min
Delta R.T.: 0.006 min
Response: 29651141
Conc: 17.28 ng/ml m



#28 Decachlorobiphenyl

R.T.: 7.844 min
Delta R.T.: -0.005 min
Response: 34853424
Conc: 15.43 ng/ml